



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113287

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/01/2021	10:00
2	CAL2	1	Water	NA	NA	20.2	7.00	03/01/2021	10:01
3	CAL3	1	Water	NA	NA	20.2	10.03	03/01/2021	10:02
4	ICV	1	Water	NA	NA	20.3	7.01	03/01/2021	10:05
5	CCV1	1	Water	NA	NA	20.2	2.01	03/01/2021	10:06
6	M1507-04	1	Water	NA	NA	20.9	7.16	03/01/2021	10:08
7	M1507-04DUP	1	Water	NA	NA	20.8	7.17	03/01/2021	10:09
8	M1507-05	1	Water	NA	NA	20.8	7.22	03/01/2021	10:10
9	M1507-07	1	Water	NA	NA	20.9	5.54	03/01/2021	10:12
10	M1507-08	1	Water	NA	NA	21.0	5.29	03/01/2021	10:15
11	M1507-09	1	Water	NA	NA	21.0	5.40	03/01/2021	10:16
12	M1508-05	1	Water	NA	NA	20.5	6.14	03/01/2021	10:19
13	CCV2	1	Water	NA	NA	20.3	12.02	03/01/2021	10:20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-water-022621 WorkList ID : 146715 Department : Wet-Chemistry Date : 02-26-2021 15:35:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/26/2021	Water	M1507-04	pH	Cool 4 deg C	PSEG03	I33	41114	02/26/2021	9040C
02/26/2021	Water	M1507-05	pH	Cool 4 deg C	PSEG03	I33	41115	02/26/2021	9040C
02/26/2021	Water	M1507-07	pH	Cool 4 deg C	PSEG03	I33	COMP-30771-73-77	02/26/2021	9040C
02/26/2021	Water	M1507-08	pH	Cool 4 deg C	PSEG03	I33	41116	02/26/2021	9040C
02/26/2021	Water	M1507-09	pH	Cool 4 deg C	PSEG03	I33	30769	02/26/2021	9040C
02/26/2021	Water	M1508-05	pH	Cool 4 deg C	PSEG03	I33	TRE-1159-1160	02/26/2021	9040C

Date/Time: 02/26/2021 8:00 AM
 Received by: SD West
 Relinquished by: AM (wet-chem)

Date/Time: 03/01/2021 11:00
 Received by: AM (wet-chem)
 Relinquished by: SD West