



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113288

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/01/2021	10:00
2	CAL2	1	Water	NA	NA	20.2	7.00	03/01/2021	10:01
3	CAL3	1	Water	NA	NA	20.2	10.03	03/01/2021	10:02
4	ICV	1	Water	NA	NA	20.3	7.01	03/01/2021	10:05
5	CCV1	1	Water	NA	NA	20.2	2.01	03/01/2021	10:06
6	M1508-02	1	Solid	20.03	20	20.7	7.30	03/01/2021	10:25
7	M1508-02DUP	1	Solid	20.04	20	20.8	7.31	03/01/2021	10:26
8	M1509-01	1	Solid	20.02	20	20.1	7.29	03/01/2021	10:29
9	CCV2	1	Water	NA	NA	20.3	12.00	03/01/2021	10:30

WORKLIST(Hardcopy Internal Chain)

L813288

WorkList Name : pH-soil-022621-2 WorkList ID : 146717 Department : Wet-Chemistry Date : 02-26-2021 15:35:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/26/2021	Solid	M1508-02	pH	Cool 4 deg C	PSEG03	I33	TRE-1138	02/26/2021	9045D
02/26/2021	Solid	M1509-01	pH	Cool 4 deg C	PSEG05	I33	BU-02-022621	02/26/2021	9045D

Date/Time 03/01/2021 8:00
 Received by: SD Ceed
 Relinquished by: AM Wetchem

Date/Time 03/01/2020 11:00
 Received by: AM Wetchem
 Relinquished by: SD Ceed