

7470A


LB11329J

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40973



A= 0.0000e+000

B= 2.4359e-004

C= 1.3418e-003

Rho= 0.9999177

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.003	-0.003	-18	0.000	-18					0/00
0.2	0.200	0.145	-0.055	589	0.0 %	589					-28
2.5	2.500	2.538	0.038	10413	0.0 %	10413					2
5.0	5.000	5.082	0.082	20857	0.0 %	20857					2
7.5	7.500	7.456	-0.044	30604	0.0 %	30604					-1
10.0	10.000	9.982	-0.018	40973	0.0 %	40973					0

LB113295 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Mar 2021 10:19:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	-18	-	0.0000	7470A	PPB	01 Mar 2021 10:22:16	S	Std
0.2 - 1	SD	589	-	0.2000	7470A	PPB	01 Mar 2021 10:24:08	S	Std
2.5 - 1	SD	10413	-	2.5000	7470A	PPB	01 Mar 2021 10:27:04	S	Std
5.0 - 1	SD	20857	-	5.0000	7470A	PPB	01 Mar 2021 10:29:21	S	Std
7.5 - 1	SD	30604	-	7.5000	7470A	PPB	01 Mar 2021 10:33:25	S	Std
10.0 - 1	SD	40973	-	10.0000	7470A	PPB	01 Mar 2021 10:40:57	S	Std
ICV18 - 1	ICV18	16170	3.9402	-7470A	PPB	01 Mar 2021 10:43:58	U	SMPL	
ICB18 - 1	ICB18	-49	-0.0106	-7470A	PPB	01 Mar 2021 10:46:13	U	SMPL	
CCV45 - 1	CCV45	20758	5.0579	-7470A	PPB	01 Mar 2021 10:49:00	U	SMPL	
CCB45 - 1	CCB45	-45	-0.0096	-7470A	PPB	01 Mar 2021 10:51:14	U	SMPL	
CRA - 1	CRA	765	0.1877	-7470A	PPB	01 Mar 2021 11:11:59	U	SMPL	
HighStd - 1	HighStd	40623	9.8968	-7470A	PPB	01 Mar 2021 11:14:47	U	SMPL	
ChkStd - 1	ChkStd	29721	7.2412	-7470A	PPB	01 Mar 2021 11:17:02	U	SMPL	
PB134847BL - 1	PBW	-56	-0.0123	-7470A	PPB	01 Mar 2021 11:19:18	U	SMPL	
PB134847BS - 1	LCSW	16693	4.0676	-7470A	PPB	01 Mar 2021 11:21:34	U	SMPL	
M1490-01 - 1	02-25-A	27	0.0079	-7470A	PPB	01 Mar 2021 11:25:38	U	SMPL	
M1490-01DUP - 1	02-25-ADUP	31	0.0089	-7470A	PPB	01 Mar 2021 11:27:54	U	SMPL	
M1490-01MS - 1	02-25-AMS	15568	3.7936	-7470A	PPB	01 Mar 2021 11:31:56	U	SMPL	
M1490-01MSD - 1	02-25-AMSD	16522	4.0260	-7470A	PPB	01 Mar 2021 11:34:13	U	SMPL	
M1490-02 - 1	02-25-B	-20	-0.0035	-7470A	PPB	01 Mar 2021 11:36:28	U	SMPL	
CCV46 - 1	CCV46	20935	5.1010	-7470A	PPB	01 Mar 2021 11:38:43	U	SMPL	
CCB46 - 1	CCB46	-30	-0.0060	-7470A	PPB	01 Mar 2021 11:40:59	U	SMPL	
M1490-03 - 1	02-25-C	28	0.0082	-7470A	PPB	01 Mar 2021 11:43:15	U	SMPL	
M1490-04 - 1	02-25-SOIL	39	0.0108	-7470A	PPB	01 Mar 2021 11:45:30	U	SMPL	
M1503-05 - 1	SEDGWICK-AVE	43	0.0118	-7470A	PPB	01 Mar 2021 11:47:46	U	SMPL	
PB134847TB - 1	PB134836TB	37	0.0104	-7470A	PPB	01 Mar 2021 11:50:01	U	SMPL	
M1490-01LX5 - 1		34	0.0096	-7470A	PPB	01 Mar 2021 11:52:17	U	SMPL	
M1490-01A - 1		16117	3.9273	-7470A	PPB	01 Mar 2021 11:54:34	U	SMPL	
CCV47 - 1	CCV47	20785	5.0644	-7470A	PPB	01 Mar 2021 11:56:51	U	SMPL	
CCB47 - 1	CCB47	-28	-0.0055	-7470A	PPB	01 Mar 2021 11:59:08	U	SMPL	