



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113298

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	02/26/2021	12:15
2	M1485-04		1	9	100.00	>212.0	>212.0	02/26/2021	12:45
3	M1485-04DUP		1	9	100.00	>212.0	>212.0	02/26/2021	13:10
4	M1485-06		1	11	100.00	>212.0	>212.0	02/26/2021	13:35
5	M1486-05		1	10	100.00	>212.0	>212.0	02/26/2021	14:05
6	M1487-05		1	10	100.00	>212.0	>212.0	02/26/2021	14:32
7	M1487-07		1	11	100.00	>212.0	>212.0	03/01/2021	15:05
8	M1488-04		1	10	100.00	>212.0	>212.0	02/26/2021	15:35
9	M1488-05		1	9	100.00	>212.0	>212.0	02/26/2021	15:58
10	M1488-06		1	10	100.00	>212.0	>212.0	02/26/2021	16:25
11	M1488-07		1	11	100.00	>212.0	>212.0	02/26/2021	16:55
12	M1489-04		1	9	100.00	>212.0	>212.0	02/26/2021	17:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP22621

WorkList ID : 146710

Department : Wet-Chemistry

Date : 02-26-2021 11:38:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/04/2021	Water	M1485-04	Flash Point	Cool 4 deg C	PSEG03	I31	35-36-84	02/25/2021	1010A
03/04/2021	Water	M1485-06	Flash Point	Cool 4 deg C	PSEG03	I31	32937	02/25/2021	1010A
03/04/2021	Water	M1486-05	Flash Point	Cool 4 deg C	PSEG03	I23	56-59	02/25/2021	1010A
03/04/2021	Water	M1487-05	Flash Point	Cool 4 deg C	PSEG03	I23	16-19	02/25/2021	1010A
03/04/2021	Water	M1487-07	Flash Point	Cool 4 deg C	PSEG03	I23	32979	02/25/2021	1010A
03/04/2021	Water	M1488-04	Flash Point	Cool 4 deg C	PSEG03	I32	41709	02/25/2021	1010A
03/04/2021	Water	M1488-05	Flash Point	Cool 4 deg C	PSEG03	I32	41698	02/25/2021	1010A
03/04/2021	Water	M1488-06	Flash Point	Cool 4 deg C	PSEG03	I32	41697	02/25/2021	1010A
03/04/2021	Water	M1488-07	Flash Point	Cool 4 deg C	PSEG03	I32	30779	02/25/2021	1010A
03/04/2021	Water	M1489-04	Flash Point	Cool 4 deg C	PSEG03	I33	COMP-41071-72	02/25/2021	1010A

Date/Time 02/26/21 11:45
 Received by: KS (Wetchem)
 Relinquished by: RC C-x+(19b)

Date/Time 02/26/21 17:25
 Received by: RC C-x+(19b)
 Relinquished by: KS (Wetchem)