



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113299

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	03/01/2021	12:20
2	M1489-05		1	9	100.00	>212.0	>212.0	03/01/2021	12:45
3	M1489-05DUP		1	9	100.00	>212.0	>212.0	03/01/2021	13:10
4	M1489-06		1	10	100.00	>212.0	>212.0	03/01/2021	13:35
5	M1507-06		1	11	100.00	>212.0	>212.0	03/01/2021	13:58
6	M1507-07		1	10	100.00	>212.0	>212.0	03/01/2021	14:25
7	M1507-08		1	11	100.00	>212.0	>212.0	03/01/2021	14:55
8	M1507-09		1	10	100.00	>212.0	>212.0	03/01/2021	15:15
9	M1508-05		1	11	100.00	>212.0	>212.0	03/01/2021	15:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB113299

WorkList Name : fp3121

WorkList ID : 146743

Department : Wet-Chemistry

Date : 03-01-2021 11:54:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/04/2021	Water	M1489-05	Flash Point	Cool 4 deg C	PSEG03	I33	COMP-41706-07	02/25/2021	1010A
03/04/2021	Water	M1489-06	Flash Point	Cool 4 deg C	PSEG03	I33	41129	02/25/2021	1010A
03/05/2021	Water	M1507-06	Flash Point	Cool 4 deg C	PSEG03	I33	COMP-14-15	02/26/2021	1010A
03/05/2021	Water	M1507-07	Flash Point	Cool 4 deg C	PSEG03	I33	COMP-30771-73-77	02/26/2021	1010A
03/05/2021	Water	M1507-08	Flash Point	Cool 4 deg C	PSEG03	I33	41116	02/26/2021	1010A
03/05/2021	Water	M1507-09	Flash Point	Cool 4 deg C	PSEG03	I33	30769	02/26/2021	1010A
03/12/2021	Water	M1508-05	Flash Point	Cool 4 deg C	PSEG03	I33	TRE-1159-1160	02/26/2021	1010A

Date/Time 03/01/21 11:55
 Received by: KS (Wetchoy)
 Relinquished by: REC Ext (96)

Date/Time 03/01/21 15:45
 Received by: REC Ext (96)
 Relinquished by: KS (Wetchoy)