



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113302

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/01/2021	13:13
2	CAL2	1	Water	NA	NA	20.2	7.00	03/01/2021	13:14
3	CAL3	1	Water	NA	NA	20.2	10.03	03/01/2021	13:15
4	ICV	1	Water	NA	NA	20.2	6.99	03/01/2021	13:16
5	CCV1	1	Water	NA	NA	21.2	2.01	03/01/2021	13:19
6	M1524-01	1	Water	NA	NA	20.2	4.91	03/01/2021	13:20
7	M1524-01DUP	1	Water	NA	NA	20.3	4.92	03/01/2021	13:21
8	CCV2	1	Water	NA	NA	20.2	12.02	03/01/2021	13:22

2B 113302

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1524 WorkList ID : 146751 Department : Wet-Chemistry Date : 03-01-2021 13:02:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/24/2021	Water	M1524-01-1	pH	Cool 4 deg C	STER08	I11	POOL-WATER	02/24/2021	9040C

Date/Time 03-01-2021 13:11:10
 Received by: JB Cook Open
 Relinquished by: JB Cook Open

Date/Time 03-01-2021 14:11:00
 Received by: JB Cook Open
 Relinquished by: JB Cook Open