


LB113303

ISM02.4

INSTRUMENT ID: CV1

 μ Abs.:

44631



A= 0.0000e+000

B= 2.2447e-004 *slope*C= -1.3759e-002 *y-intercept*

Rho= 0.9999475

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.00	0.000	0.001	0.001	65	0.000	65					=
0.05	0.050					332					-2
0.20	0.200	0.196	-0.004	933	0.0 %	933					-1
2.50	2.500	2.472	-0.028	11074	0.0 %	11074					-1
5.00	5.000	5.074	0.074	22666	0.0 %	22666					-1
7.50	7.500	7.452	-0.048	33261	0.0 %	33261					-1
10.0	10.000	10.005	0.005	44631	0.0 %	44631					0

LB113303
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	65	-		0 ISM02.4	PPB	3/1/2021 13:57	Std
	0.2 S0.2	933	-		0.2 ISM02.4	PPB	3/1/2021 13:59	Std
	2.5 S2.5	11074	-		2.5 ISM02.4	PPB	3/1/2021 14:01	Std
	5 S5.0	22666	-		5 ISM02.4	PPB	3/1/2021 14:04	Std
	7.5 S7.5	33261	-		7.5 ISM02.4	PPB	3/1/2021 14:06	Std
	10 S10.0	44631	-		10 ISM02.4	PPB	3/1/2021 14:11	Std
ICV	ICV	17392	3.8903	-	ISM02.4	PPB	3/1/2021 14:13	SMPL
ICB	ICB	1	-0.0135	-	ISM02.4	PPB	3/1/2021 14:16	SMPL
CCV048	CCV048	21838	4.8883	-	ISM02.4	PPB	3/1/2021 14:18	SMPL
CCB048	CCB048	-8	-0.0156	-	ISM02.4	PPB	3/1/2021 14:20	SMPL
PB134883BL	PBW883	27	-0.0077	-	ISM02.4	PPB	3/1/2021 14:22	SMPL
M1512-01	METZC4	40	-0.0048	-	ISM02.4	PPB	3/1/2021 14:25	SMPL
M1512-02	METZC4D	37	-0.0055	-	ISM02.4	PPB	3/1/2021 14:27	SMPL
M1512-03	METZC4S	4675	1.0357	-	ISM02.4	PPB	3/1/2021 14:29	SMPL
M1512-04	METZC5	27	-0.0077	-	ISM02.4	PPB	3/1/2021 14:31	SMPL
M1512-05	METZC6	27	-0.0077	-	ISM02.4	PPB	3/1/2021 14:34	SMPL
M1512-06	METZC7	57	-0.001	-	ISM02.4	PPB	3/1/2021 14:36	SMPL
M1512-07	METZC8	58	-0.0007	-	ISM02.4	PPB	3/1/2021 14:38	SMPL
CCV049	CCV049	21649	4.8459	-	ISM02.4	PPB	3/1/2021 14:41	SMPL
CCB049	CCB049	2	-0.0133	-	ISM02.4	PPB	3/1/2021 14:43	SMPL