



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 3/2/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 03/01/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:05
Out Date: 03/02/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113306

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M1458-16	LLOQ-SOIL-QT1-2021	1	1.00	2.00	2.00	100.0	
M1525-01	OK-01-030121	32	1.15	9.61	8.76	90.0	
M1525-02	OK-01-030121-EPH-2	33	1.16	9.81	8.97	90.3	
M1526-01	STOCK-PILE	2	1.13	9.78	8.35	83.5	
M1526-02	STOCK-PILE-EPH-2	3	1.14	9.8	8.39	83.7	
M1527-01	ETGI-371	34	1.16	9.96	8.62	84.8	
M1527-02	ETGI-371	35	1.1	9.8	8.61	86.3	
M1528-01	MW-CON	4	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M1529-01	KMA754C-1-1	5	1.00	2.00	2.00	100.0	PILC REEL
M1529-02	KMA754C-1-2	6	1.00	2.00	2.00	100.0	PILC REEL
M1530-01	HEF265P-1-1	7	1.00	2.00	2.00	100.0	PILC REEL
M1530-02	HEF265P-2-1	8	1.00	2.00	2.00	100.0	PILC REEL
M1530-03	HEF265P-2-2	9	1.00	2.00	2.00	100.0	PILC REEL
M1530-04	HEF265P-1-2	10	1.00	2.00	2.00	100.0	PILC REEL
M1530-05	HEE442S-1-1	11	1.00	2.00	2.00	100.0	PILC REEL
M1530-06	HEE442S-1-2	12	1.00	2.00	2.00	100.0	PILC REEL
M1530-07	HEE442S-1-3	13	1.00	2.00	2.00	100.0	PILC REEL
M1531-01	HO-1	14	1.16	9.96	9.04	89.5	
M1532-01	SOIL-SAMPLE-S-1	15	1.12	9.76	8.75	88.3	
M1532-02	SOIL-SAMPLE-S-2	16	1.16	9.67	8.86	90.5	
M1532-02D	SOIL-SAMPLE-S-2DUP	17	1.18	9.69	8.88	90.5	
M1534-08	MDL-SOIL-01	18	1.00	2.00	2.00	100.0	
M1534-09	MDL-SOIL-02	19	1.00	2.00	2.00	100.0	
M1534-10	MDL-SOIL-03	20	1.00	2.00	2.00	100.0	
M1534-11	MDL-SOIL-04	21	1.00	2.00	2.00	100.0	
M1534-12	MDL-SOIL-05	22	1.00	2.00	2.00	100.0	
M1534-13	MDL-SOIL-06	23	1.00	2.00	2.00	100.0	
M1534-14	MDL-SOIL-07	24	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 3/2/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 03/01/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:05
Out Date: 03/02/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113306

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M1534-15	MDL-ME-SOIL-01	25	1.00	2.00	2.00	100.0	
M1534-16	MDL-ME-SOIL-02	26	1.00	2.00	2.00	100.0	
M1534-17	MDL-ME-SOIL-03	27	1.00	2.00	2.00	100.0	
M1534-18	MDL-ME-SOIL-04	28	1.00	2.00	2.00	100.0	
M1534-19	MDL-ME-SOIL-05	29	1.00	2.00	2.00	100.0	
M1534-20	MDL-ME-SOIL-06	30	1.00	2.00	2.00	100.0	
M1534-21	MDL-ME-SOIL-07	31	1.00	2.00	2.00	100.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$