


LB113307

7471B

INSTRUMENT ID: CV1

μ Abs.:

55065



Linear

A= 0.0000e+000

B= 1.7962e-004

C= 3.0904e-002

Rho= 0.9997293

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	0.041	0.041	58	0.000	58					100
0.2	0.200	0.219	0.019	1049	0.0 %	1049					10
2.5	2.500	2.356	-0.144	12943	0.0 %	12943					-6
5.0	5.000	5.048	0.048	27929	0.0 %	27929					1
7.5	7.500	7.614	0.114	42216	0.0 %	42216					2
10.0	10.000	9.922	-0.078	55065	0.0 %	55065					-1

LB113307 INSTRUMENT ID :: CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Mar 2021 15:58:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.2	58	-	0.00007471B	PPB		01 Mar 2021 15:59:02	S	Std
0.2 - 1	52.5	1049	-	0.20007471B	PPB		01 Mar 2021 16:01:20	S	Std
2.5 - 1	55	12943	-	2.50007471B	PPB		01 Mar 2021 16:03:37	S	Std
5.0 - 1	57.5	27929	-	5.00007471B	PPB		01 Mar 2021 16:05:54	S	Std
7.5 - 1	58	42216	-	7.50007471B	PPB		01 Mar 2021 16:08:11	S	Std
10.0 - 1	58	55065	-	10.00007471B	PPB		01 Mar 2021 16:10:27	S	Std
ICV19 - 1	ICV19	21358	3.8673	-7471B	PPB		01 Mar 2021 16:13:08	U	SMPL
ICB19 - 1	ICB19	-51	0.0217	-7471B	PPB		01 Mar 2021 16:15:23	U	SMPL
CCV50 - 1	CCV50	27935	5.0487	-7471B	PPB		01 Mar 2021 16:17:38	U	SMPL
CCB50 - 1	CCB50	-50	0.0219	-7471B	PPB		01 Mar 2021 16:19:53	U	SMPL
CRA - 1	CRA	1038	0.2174	-7471B	PPB		01 Mar 2021 16:22:08	U	SMPL
HighStd - 1	HighStd	55181	9.9428	-7471B	PPB		01 Mar 2021 16:24:23	U	SMPL
ChkStd - 1	ChkStd	40097	7.2333	-7471B	PPB		01 Mar 2021 16:26:38	U	SMPL
PB134884BL - 1	PBS	-91	0.0146	-7471B	PPB		01 Mar 2021 16:28:55	U	SMPL
PB134884BS - 1	LCSS	21376	3.8706	-7471B	PPB		01 Mar 2021 16:31:12	U	SMPL
M1489-03 - 1	40030	252	0.0762	-7471B	PPB		01 Mar 2021 16:33:28	U	SMPL
M1489-11 - 1	41129	164	0.0604	-7471B	PPB		01 Mar 2021 16:35:45	U	SMPL
M1506-01 - 1	OR-03-022621	784	0.1717	-7471B	PPB		01 Mar 2021 16:38:02	U	SMPL
M1507-03 - 1	41007	625	0.1432	-7471B	PPB		01 Mar 2021 16:40:19	U	SMPL
CCV51 - 1	CCV51	27939	5.0494	-7471B	PPB		01 Mar 2021 16:42:35	U	SMPL
CCB51 - 1	CCB51	-58	0.0205	-7471B	PPB		01 Mar 2021 16:44:50	U	SMPL
M1509-01 - 1	BU-02-022621	1740	0.3435	-7471B	PPB		01 Mar 2021 16:47:05	U	SMPL
M1509-01DUP - 1	BU-02-022621DUP	1646	0.3266	-7471B	PPB		01 Mar 2021 16:49:21	U	SMPL
M1509-01MS - 1	BU-02-022621MS	23567	4.2641	-7471B	PPB		01 Mar 2021 16:51:36	U	SMPL
M1509-01MSD - 1	BU-02-022621MSD	23727	4.2929	-7471B	PPB		01 Mar 2021 16:53:52	U	SMPL
M1509-01LX5 - 1		259	0.0774	-7471B	PPB		01 Mar 2021 16:56:08	U	SMPL
M1509-01A - 1		25262	4.5686	-7471B	PPB		01 Mar 2021 16:58:24	U	SMPL
CCV52 - 1	CCV52	27941	5.0498	-7471B	PPB		01 Mar 2021 17:00:40	U	SMPL
CCB52 - 1	CCB52	-40	0.0237	-7471B	PPB		01 Mar 2021 17:02:57	U	SMPL