

LB113312


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40221



A= 0.0000e+000

B= 2.4713e-004

C= -1.8894e-003

Rho= 0.9999172

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.006	0.006	31	0.000	31					0
0.2	0.200	0.200	0.000	818	0.0 %	818					-2
2.5	2.500	2.450	-0.050	9922	0.0 %	9922					1
5.0	5.000	5.025	0.025	20341	0.0 %	20341					1
7.5	7.500	7.581	0.081	30683	0.0 %	30683					1
10.0	10.000	9.938	-0.062	40221	0.0 %	40221					1

LB113318 INSTRUMENT ID : CV : 1

Method: 7470A		Operator: Admin		Date of Analysis: 02 Mar 2021 10:42:58					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	7470A	PPB	02 Mar 2021 10:49:14	S	Std
0.2 - 1	50.2	818	-	0.2000	7470A	PPB	02 Mar 2021 10:51:32	S	Std
2.5 - 1	52.5	9922	-	2.5000	7470A	PPB	02 Mar 2021 10:53:49	S	Std
5.0 - 1	55	20341	-	5.0000	7470A	PPB	02 Mar 2021 10:56:06	S	Std
7.5 - 1	57.5	30683	-	7.5000	7470A	PPB	02 Mar 2021 10:58:23	S	Std
10.0 - 1	80	40221	-	10.0000	7470A	PPB	02 Mar 2021 11:06:08	S	Std
ICV20 - 1	ICV20	15204	3.7555	-	7470A	PPB	02 Mar 2021 11:09:19	U	SMPL
ICB20 - 1	ICB20	-34	-0.0103	-	7470A	PPB	02 Mar 2021 11:11:34	U	SMPL
CCV53 - 1	CCV53	19693	4.8649	-	7470A	PPB	02 Mar 2021 11:13:49	U	SMPL
CCB53 - 1	CCB53	-37	-0.0110	-	7470A	PPB	02 Mar 2021 11:16:04	U	SMPL
CRA - 1	CRA	790	0.1933	-	7470A	PPB	02 Mar 2021 11:18:19	U	SMPL
HighStd - 1	HighStd	39662	9.7998	-	7470A	PPB	02 Mar 2021 11:20:34	U	SMPL
ChkStd - 1	ChkStd	29282	7.2346	-	7470A	PPB	02 Mar 2021 11:22:49	U	SMPL
PB134885BL - 1	PBW	-68	-0.0187	-	7470A	PPB	02 Mar 2021 11:25:06	U	SMPL
PB134885BS - 1	LCSW	15538	3.8380	-	7470A	PPB	02 Mar 2021 11:27:21	U	SMPL
M1488-01 - 1	41700	-5	-0.0031	-	7470A	PPB	02 Mar 2021 11:29:38	U	SMPL
M1488-01DUP - 1	41700DUP	23	0.0038	-	7470A	PPB	02 Mar 2021 11:31:54	U	SMPL
M1488-01MS - 1	41700MS	16019	3.9569	-	7470A	PPB	02 Mar 2021 11:34:12	U	SMPL
M1488-01MSD - 1	41700MSD	15419	3.8086	-	7470A	PPB	02 Mar 2021 11:36:29	U	SMPL
M1488-03 - 1	41701	6	-0.0004	-	7470A	PPB	02 Mar 2021 11:38:45	U	SMPL
CCV54 - 1	CCV54	19914	4.9195	-	7470A	PPB	02 Mar 2021 11:41:00	U	SMPL
CCB54 - 1	CCB54	-29	-0.0091	-	7470A	PPB	02 Mar 2021 11:43:15	U	SMPL
M1489-01 - 1	COMP-41087-88	16	0.0021	-	7470A	PPB	02 Mar 2021 11:45:31	U	SMPL
M1508-01 - 1	TRE-1138	24	0.0040	-	7470A	PPB	02 Mar 2021 11:47:47	U	SMPL
PB134886BL - 1	PBW	33	0.0063	-	7470A	PPB	02 Mar 2021 11:50:03	U	SMPL
PB134886BS - 1	LCSW	15990	3.9497	-	7470A	PPB	02 Mar 2021 11:52:18	U	SMPL
M1489-07 - 1	41108	86	0.0194	-	7470A	PPB	02 Mar 2021 11:54:34	U	SMPL
M1508-04 - 1	TRE-1159-1160	7	-0.0002	-	7470A	PPB	02 Mar 2021 11:56:51	U	SMPL
M1508-04DUP - 1	TRE-1159-1160DUP	25	0.0043	-	7470A	PPB	02 Mar 2021 11:59:07	U	SMPL
M1508-04MS - 1	TRE-1159-1160MS	15907	3.9292	-	7470A	PPB	02 Mar 2021 12:01:24	U	SMPL
M1508-04MSD - 1	TRE-1159-1160MSD	16361	4.0414	-	7470A	PPB	02 Mar 2021 12:03:42	U	SMPL
M1517-08 - 1	INK-PIT	52	0.0110	-	7470A	PPB	02 Mar 2021 12:05:57	U	SMPL
CCV55 - 1	CCV55	20005	4.9420	-	7470A	PPB	02 Mar 2021 12:08:13	U	SMPL
CCB55 - 1	CCB55	-31	-0.0096	-	7470A	PPB	02 Mar 2021 12:10:29	U	SMPL
M1517-09 - 1	INK-TOTE	314	0.0757	-	7470A	PPB	02 Mar 2021 12:12:44	U	SMPL
PB134887BL - 1	PBW	-19	-0.0066	-	7470A	PPB	02 Mar 2021 12:15:00	U	SMPL
PB134887BS - 1	LCSW	17366	4.2898	-	7470A	PPB	02 Mar 2021 12:17:16	U	SMPL
M1489-06 - 1	41129	405	0.0982	-	7470A	PPB	02 Mar 2021 12:19:32	U	SMPL
M1507-06 - 1	COMP-14-15	741	0.1812	-	7470A	PPB	02 Mar 2021 12:21:47	U	SMPL
M1507-06DUP - 1	COMP-14-15DUP	810	0.1983	-	7470A	PPB	02 Mar 2021 12:24:04	U	SMPL
M1507-06MS - 1	COMP-14-15MS	17001	4.1996	-	7470A	PPB	02 Mar 2021 12:26:20	U	SMPL
M1507-06MSD - 1	COMP-14-15MSD	17197	4.2480	-	7470A	PPB	02 Mar 2021 12:28:37	U	SMPL
M1517-06 - 1	INK-PIT	157	0.0369	-	7470A	PPB	02 Mar 2021 12:30:54	U	SMPL
M1517-07 - 1	INK-TOTE	607	0.1481	-	7470A	PPB	02 Mar 2021 12:33:10	U	SMPL
CCV56 - 1	CCV56	19987	4.9375	-	7470A	PPB	02 Mar 2021 12:35:25	U	SMPL
CCB56 - 1	CCB56	-41	-0.0120	-	7470A	PPB	02 Mar 2021 12:37:41	U	SMPL
PB134885TB - 1	PB134885TB	25	0.0043	-	7470A	PPB	02 Mar 2021 12:39:57	U	SMPL
PB134886TB - 1	PB134886TB	39	0.0077	-	7470A	PPB	02 Mar 2021 12:42:12	U	SMPL
M1488-01LX5 - 1		33	0.0063	-	7470A	PPB	02 Mar 2021 12:44:28	U	SMPL
M1488-01A - 1		15247	3.7661	-	7470A	PPB	02 Mar 2021 12:46:44	U	SMPL
M1508-04LX5 - 1		-19	-0.0066	-	7470A	PPB	02 Mar 2021 12:49:00	U	SMPL
M1508-04A - 1		15306	3.7807	-	7470A	PPB	02 Mar 2021 12:51:17	U	SMPL
M1507-06LX5 - 1		111	0.0255	-	7470A	PPB	02 Mar 2021 12:53:33	U	SMPL
M1507-06A - 1		16959	4.1892	-	7470A	PPB	02 Mar 2021 12:55:51	U	SMPL
CCV57 - 1	CCV57	19421	4.7976	-	7470A	PPB	02 Mar 2021 12:58:07	U	SMPL
CCB57 - 1	CCB57	-58	-0.0162	-	7470A	PPB	02 Mar 2021 13:00:23	U	SMPL