


LB113321

7471B

INSTRUMENT ID: CV1

 μ Abs.:

54109



Linear

A= 0.0000e+000

B= 1.8241e-004

C= -3.4556e-002

Rho= 0.9995721

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.011	-0.011	130	0.000	130					8/1/17
0.2	0.200	0.179	-0.021	1173	0.0 %	1173					-11
2.5	2.500	2.433	-0.067	13528	0.0 %	13528					-3
5.0	5.000	5.090	0.090	28093	0.0 %	28093					2
7.5	7.500	7.673	0.173	42251	0.0 %	42251					2
10.0	10.000	9.836	-0.164	54109	0.0 %	54109					-2

LB113321 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 02 Mar 2021 14:29:58					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	130	-	0.0000	7471B	PPB	02 Mar 2021 14:49:18	S	Std
0.2 - 1	50.2	1173	-	0.2000	7471B	PPB	02 Mar 2021 14:51:34	S	Std
2.5 - 1	52.5	13528	-	2.5000	7471B	PPB	02 Mar 2021 14:53:51	S	Std
5.0 - 1	55	28093	-	5.0000	7471B	PPB	02 Mar 2021 14:56:08	S	Std
7.5 - 1	57.5	42251	-	7.5000	7471B	PPB	02 Mar 2021 14:58:24	S	Std
10.0 - 1	510	54109	-	10.0000	7471B	PPB	02 Mar 2021 15:00:40	S	Std
ICV21 - 1	ICV21	21400	3.8691	-	7471B	PPB	02 Mar 2021 15:03:35	U	SMPL
ICB21 - 1	ICB21	86	-0.0189	-	7471B	PPB	02 Mar 2021 15:05:50	U	SMPL
CCV58 - 1	CCV58	27851	5.0459	-	7471B	PPB	02 Mar 2021 15:08:05	U	SMPL
CCB58 - 1	CCB58	69	-0.0220	-	7471B	PPB	02 Mar 2021 15:10:20	U	SMPL
CRA - 1	CRA	1154	0.1759	-	7471B	PPB	02 Mar 2021 15:12:35	U	SMPL
HighStd - 1	HighStd	53599	9.7426	-	7471B	PPB	02 Mar 2021 15:14:50	U	SMPL
ChkStd - 1	ChkStd	39334	7.1405	-	7471B	PPB	02 Mar 2021 15:17:05	U	SMPL
PB134909BL - 1	PBS	42	-0.0269	-	7471B	PPB	02 Mar 2021 15:19:22	U	SMPL
PB134909BS - 1	LCSS	24810	4.4911	-	7471B	PPB	02 Mar 2021 15:21:39	U	SMPL
M1525-01 - 1	OK-01-030121	1227	0.1893	-	7471B	PPB	02 Mar 2021 15:23:56	U	SMPL
M1526-01 - 1	STOCK-PILE	3469	0.5982	-	7471B	PPB	02 Mar 2021 15:26:13	U	SMPL
M1527-01 - 1	ETGI-371	773	0.1064	-	7471B	PPB	02 Mar 2021 15:33:20	U	SMPL
M1527-01DUP - 1	ETGI-371DUP	844	0.1194	-	7471B	PPB	02 Mar 2021 15:35:36	U	SMPL
M1527-01MS - 1	ETGI-371MS	22381	4.0480	-	7471B	PPB	02 Mar 2021 15:37:52	U	SMPL
CCV59 - 1	CCV59	27794	5.0355	-	7471B	PPB	02 Mar 2021 15:40:07	U	SMPL
CCB59 - 1	CCB59	84	-0.0192	-	7471B	PPB	02 Mar 2021 15:42:23	U	SMPL
M1527-01MSD - 1	ETGI-371MSD	22778	4.1205	-	7471B	PPB	02 Mar 2021 15:44:38	U	SMPL
M1527-01LX5 - 1		193	0.0006	-	7471B	PPB	02 Mar 2021 15:46:53	U	SMPL
M1527-01A - 1		23368	4.2281	-	7471B	PPB	02 Mar 2021 15:49:08	U	SMPL
CCV60 - 1	CCV60	27582	4.9968	-	7471B	PPB	02 Mar 2021 15:51:24	U	SMPL
CCB60 - 1	CCB60	78	-0.0203	-	7471B	PPB	02 Mar 2021 15:53:39	U	SMPL