



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113323

Slope : 97.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	03/02/2021	15:15
2	CAL2	1	Water	NA	NA	20.1	7.00	03/02/2021	15:16
3	CAL3	1	Water	NA	NA	20.1	10.03	03/02/2021	15:17
4	ICV	1	Water	NA	NA	20.2	6.99	03/02/2021	15:19
5	CCV1	1	Water	NA	NA	20.2	2.01	03/02/2021	15:20
6	M1538-01	1	Water	NA	NA	20.3	7.69	03/02/2021	15:22
7	M1538-01DUP	1	Water	NA	NA	20.4	7.70	03/02/2021	15:23
8	M1541-02	1	Water	NA	NA	20.4	7.42	03/02/2021	15:25
9	M1545-01	1	Water	NA	NA	20.3	7.56	03/02/2021	15:26
10	M1545-02	1	Water	NA	NA	20.2	7.55	03/02/2021	15:27
11	M1545-03	1	Water	NA	NA	20.3	7.34	03/02/2021	15:30
12	M1545-06	1	Water	NA	NA	20.2	8.26	03/02/2021	15:32
13	M1545-07	1	Water	NA	NA	20.4	8.12	03/02/2021	15:34
14	M1545-08	1	Water	NA	NA	20.4	7.95	03/02/2021	15:36
15	M1545-20	1	Water	NA	NA	20.2	7.67	03/02/2021	15:39
16	CCV2	1	Water	NA	NA	20.2	12.02	03/02/2021	15:40
17	M1545-21	1	Water	NA	NA	20.1	7.81	03/02/2021	15:42
18	CCV3	1	Water	NA	NA	20.3	2.01	03/02/2021	15:45

LB 113323

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1541 WorkList ID : 146796 Department : Wet-Chemistry Date : 03-02-2021 14:58:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2021	Water	M1538-01	pH	Cool 4 deg C	PSEG03	I51	1-2-TOTES	03/02/2021	9040C
03/02/2021	Water	M1541-02	pH	Cool 4 deg C	PSEG03	I51	FM-786	03/02/2021	9040C
03/02/2021	Water	M1545-01	pH	Cool 4 deg C	TULL02	I13	MH-5	03/02/2021	9040C
03/02/2021	Water	M1545-02	pH	Cool 4 deg C	TULL02	I13	MH-725	03/02/2021	9040C
03/02/2021	Water	M1545-03	pH	Cool 4 deg C	TULL02	I13	MH-20A	03/02/2021	9040C
03/02/2021	Water	M1545-06	pH	Cool 4 deg C	TULL02	I13	MH-F	03/02/2021	9040C
03/02/2021	Water	M1545-07	pH	Cool 4 deg C	TULL02	I13	MH-J	03/02/2021	9040C
03/02/2021	Water	M1545-08	pH	Cool 4 deg C	TULL02	I13	MH-G	03/02/2021	9040C
03/02/2021	Water	M1545-20	pH	Cool 4 deg C	TULL02	I13	BNM-1	03/02/2021	9040C
03/02/2021	Water	M1545-21	pH	Cool 4 deg C	TULL02	I13	RIM-7.55	03/02/2021	9040C

Date/Time 03/02/2021 15:00
 Received by: JP (wet chem)
 Relinquished by: JP (wet chem)

Date/Time 03/02/2021 14:15
 Received by: JP (wet chem)
 Relinquished by: JP (wet chem)