



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113332

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	03/03/2021	11:35
2	CAL2	1	Water	NA	NA	21.3	7.00	03/03/2021	11:36
3	CAL3	1	Water	NA	NA	21.3	10.03	03/03/2021	11:38
4	ICV	1	Water	NA	NA	21.3	6.99	03/03/2021	11:40
5	CCV1	1	Water	NA	NA	20.2	2.01	03/03/2021	11:41
6	M1548-01	1	Water	NA	NA	20.1	7.04	03/03/2021	11:45
7	M1548-01DUP	1	Water	NA	NA	20.2	7.05	03/03/2021	11:46
8	CCV2	1	Water	NA	NA	20.3	12.02	03/03/2021	11:48

WORKLIST(Hardcopy Internal Chain)

LB 113332

WorkList Name : ph m1548

WorkList ID : 146824

Department : Wet-Chemistry

Date : 03-03-2021 11:29:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2021	Water	M1548-01	pH ✓	Cool 4 deg C	APT101	I11	WGGM-21-WW-TSF-009	03/02/2021	SM 4500-H B

Date/Time 03/03/2021 11:30

Received by: SO (No. 2)

Relinquished by: JP (Glen)

Date/Time

Received by:

Relinquished by:

03/03/2021 12:35

SO (No. 4)

JP (Clen)