

LB113333


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

54255



A= 0.0000e+000

B= 1.8401e-004

C= -4.6528e-002

Rho= 0.9999387

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.020	-0.020	143	0.000	143					-
0.2	0.200	0.195	-0.005	1311	0.0 %	1311					-3
2.5	2.500	2.484	-0.016	13750	0.0 %	13750					-1
5.0	5.000	5.051	0.051	27700	0.0 %	27700					-1
7.5	7.500	7.554	0.054	41303	0.0 %	41303					-1
10.0	10.000	9.938	-0.062	54259	0.0 %	54259					-1

LB113333 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Mar 2021 12:32:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	143	-	0.0000	7471B	PPB	03 Mar 2021 12:35:44	S	Std
0.2 - 1	50.2	1311	-	0.2000	7471B	PPB	03 Mar 2021 12:38:00	S	Std
2.5 - 1	52.5	13750	-	2.5000	7471B	PPB	03 Mar 2021 12:40:17	S	Std
5.0 - 1	55	27700	-	5.0000	7471B	PPB	03 Mar 2021 12:42:33	S	Std
7.5 - 1	57.5	41303	-	7.5000	7471B	PPB	03 Mar 2021 12:44:50	S	Std
10.0 - 1	59.5	54259	-	10.0000	7471B	PPB	03 Mar 2021 12:47:06	S	Std
ICV22 - 1	ICV22	21282	3.8696	-	7471B	PPB	03 Mar 2021 12:49:54	U	SMPL
ICB22 - 1	ICB22	-17	-0.0497	-	7471B	PPB	03 Mar 2021 12:52:09	U	SMPL
CCV61 - 1	CCV61	27882	5.0840	-	7471B	PPB	03 Mar 2021 12:54:23	U	SMPL
CCB61 - 1	CCB61	47	-0.0379	-	7471B	PPB	03 Mar 2021 12:59:09	U	SMPL
CRA - 1	CRA	1322	0.1967	-	7471B	PPB	03 Mar 2021 13:01:23	U	SMPL
HighStd - 1	HighStd	54273	9.9403	-	7471B	PPB	03 Mar 2021 13:03:38	U	SMPL
ChkStd - 1	ChkStd	38865	7.1050	-	7471B	PPB	03 Mar 2021 13:05:53	U	SMPL
PB134934BL - 1	PBS	143	-0.0202	-	7471B	PPB	03 Mar 2021 13:08:09	U	SMPL
PB134934BS - 1	LCSS	21395	3.8904	-	7471B	PPB	03 Mar 2021 13:13:02	U	SMPL
M1458-01 - 1	LOD-MDL-SOIL-01-QT1-2021	501	0.0457	-	7471B	PPB	03 Mar 2021 13:17:57	U	SMPL
M1458-02 - 1	LOQ-SOIL-02-QT1-2021	1340	0.2000	-	7471B	PPB	03 Mar 2021 13:20:13	U	SMPL
M1537-01 - 1	BU-01-030221	1370	0.2056	-	7471B	PPB	03 Mar 2021 13:22:29	U	SMPL
M1537-01DUP - 1	BU-01-030221DUP	1213	0.1767	-	7471B	PPB	03 Mar 2021 13:24:46	U	SMPL
M1537-01MS - 1	BU-01-030221MS	24162	4.3995	-	7471B	PPB	03 Mar 2021 13:29:32	U	SMPL
CCV62 - 1	CCV62	28088	5.1220	-	7471B	PPB	03 Mar 2021 13:31:47	U	SMPL
CCB62 - 1	CCB62	-4	-0.0473	-	7471B	PPB	03 Mar 2021 13:34:02	U	SMPL
M1537-01MSD - 1	BU-01-030221MSD	24223	4.4108	-	7471B	PPB	03 Mar 2021 13:36:17	U	SMPL
M1539-01 - 1	JC-02-030221	6086	1.0734	-	7471B	PPB	03 Mar 2021 13:38:33	U	SMPL
M1540-01 - 1	PR-01-030221	8818	1.5761	-	7471B	PPB	03 Mar 2021 13:40:48	U	SMPL
M1542-01 - 1	SU-03-030221	4278	0.7407	-	7471B	PPB	03 Mar 2021 13:43:03	U	SMPL
M1544-03 - 1	TT-040-IDWSO-20210301	325	0.0133	-	7471B	PPB	03 Mar 2021 13:45:19	U	SMPL
M1537-01LX5 - 1		302	0.0090	-	7471B	PPB	03 Mar 2021 13:47:36	U	SMPL
M1537-01A - 1		23014	4.1883	-	7471B	PPB	03 Mar 2021 13:57:36	U	SMPL
CCV63 - 1	CCV63	28027	5.1107	-	7471B	PPB	03 Mar 2021 13:59:52	U	SMPL
CCB63 - 1	CCB63	30	-0.0410	-	7471B	PPB	03 Mar 2021 14:02:09	U	SMPL