


LB113337

245.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4639e-004

C= -1.0659e-001

Rho= 0.9995946

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.049	-0.049	234	0.000	234					-
0.05	0.050	0.053	0.003	647	0.0 %	647	0				6
0.2	0.200	0.235	0.035	1386	0.0 %	1386					18
2.5	2.500	2.500	0.000	10579	0.0 %	10579					0
5.0	5.000	5.147	0.147	21323	0.0 %	21323					3
7.5	7.500	7.284	-0.216	29996	0.0 %	29996					-3
10.0	10.000	10.080	0.080	41341	0.0 %	41341					1

LB113337 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Mar 2021 14:44:09

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc/Method	Units	Date	Type	Type
0.0 - 1	SD	234	-	0.0000245.1	PPB	03 Mar 2021 15:01:54	S	Std
0.05 - 1	SD.2	647	-	0.0500245.1	PPB	03 Mar 2021 15:15:13	S	Std
0.2 - 1	SD.5	1386	-	0.2000245.1	PPB	03 Mar 2021 15:17:30	S	Std
2.5 - 1	SD.5	10579	-	2.5000245.1	PPB	03 Mar 2021 15:19:46	S	Std
5.0 - 1	SD.5	21323	-	5.0000245.1	PPB	03 Mar 2021 15:22:03	S	Std
7.5 - 1	SD.5	29996	-	7.5000245.1	PPB	03 Mar 2021 15:24:18	S	Std
10.0 - 1	SD	41341	-	10.0000245.1	PPB	03 Mar 2021 15:26:34	S	Std
ICV23 - 1	ICV23	16292	3.9077	-245.1	PPB	03 Mar 2021 15:32:15	U	SMPL
ICB23 - 1	ICB23	367	-0.0162	-245.1	PPB	03 Mar 2021 15:37:04	U	SMPL
CCV64 - 1	CCV64	21053	5.0808	-245.1	PPB	03 Mar 2021 15:39:19	U	SMPL
CCB64 - 1	CCB64	267	-0.0408	-245.1	PPB	03 Mar 2021 15:41:34	U	SMPL
CRA - 1	CRA	629	0.0484	-245.1	PPB	03 Mar 2021 15:43:49	U	SMPL
HighStd - 1	HighStd	42003	10.2427	-245.1	PPB	03 Mar 2021 15:46:03	U	SMPL
ChkStd - 1	ChkStd	29181	7.0835	-245.1	PPB	03 Mar 2021 15:48:19	U	SMPL
PB134928BL - 1	PBW	294	-0.0342	-245.1	PPB	03 Mar 2021 15:50:34	U	SMPL
PB134928BS - 1	LCSW	14749	3.5275	-245.1	PPB	03 Mar 2021 15:52:51	U	SMPL
M1458-10 - 1	LOD-MDL-WATER-04-QT1-2021	618	0.0457	-245.1	PPB	03 Mar 2021 16:29:22	U	SMPL
M1458-11 - 1	LOQ-WATER-05-QT1-2021	1423	0.2440	-245.1	PPB	03 Mar 2021 16:31:37	U	SMPL
M1510-01 - 1	CITY-WATER	596	0.0403	-245.1	PPB	03 Mar 2021 16:33:54	U	SMPL
M1520-02 - 1	NCQ-632	815	0.0942	-245.1	PPB	03 Mar 2021 16:36:11	U	SMPL
M1521-05 - 1	OB-722	427	-0.0014	-245.1	PPB	03 Mar 2021 16:38:26	U	SMPL
CCV65 - 1	CCV65	21238	5.1263	-245.1	PPB	03 Mar 2021 16:40:41	U	SMPL
CCB65 - 1	CCB65	311	-0.0300	-245.1	PPB	03 Mar 2021 16:45:30	U	SMPL
M1545-01 - 1	MH-5	445	0.0031	-245.1	PPB	03 Mar 2021 16:47:45	U	SMPL
M1545-01DUP - 1	MH-5DUP	203	-0.0566	-245.1	PPB	03 Mar 2021 16:50:01	U	SMPL
M1545-01MS - 1	MH-5MS	14214	3.3957	-245.1	PPB	03 Mar 2021 16:52:16	U	SMPL
M1545-01MSD - 1	MH-5MSD	14217	3.3964	-245.1	PPB	03 Mar 2021 16:54:32	U	SMPL
M1545-02 - 1	MH-725	333	-0.0245	-245.1	PPB	03 Mar 2021 16:56:47	U	SMPL
M1545-03 - 1	MH-20A	1267	0.2056	-245.1	PPB	03 Mar 2021 16:59:03	U	SMPL
M1545-06 - 1	MH-F	429	-0.0009	-245.1	PPB	03 Mar 2021 17:01:19	U	SMPL
M1545-07 - 1	MH-J	1597	0.2869	-245.1	PPB	03 Mar 2021 17:03:36	U	SMPL
M1545-08 - 1	MH-G	2451	0.4973	-245.1	PPB	03 Mar 2021 17:05:53	U	SMPL
M1545-09 - 1	MH-5	196	-0.0583	-245.1	PPB	03 Mar 2021 17:08:09	U	SMPL
CCV66 - 1	CCV66	21170	5.1096	-245.1	PPB	03 Mar 2021 17:10:24	U	SMPL
CCB66 - 1	CCB66	385	-0.0117	-245.1	PPB	03 Mar 2021 17:12:40	U	SMPL
M1545-10 - 1	MH-725	222	-0.0519	-245.1	PPB	03 Mar 2021 17:14:55	U	SMPL
M1545-11 - 1	MH-20A	499	0.0164	-245.1	PPB	03 Mar 2021 17:17:10	U	SMPL
M1545-12 - 1	MH-F	167	-0.0654	-245.1	PPB	03 Mar 2021 17:19:26	U	SMPL
M1545-13 - 1	MH-J	342	-0.0223	-245.1	PPB	03 Mar 2021 17:21:41	U	SMPL
M1545-14 - 1	MH-G	383	-0.0122	-245.1	PPB	03 Mar 2021 17:23:57	U	SMPL
M1545-20 - 1	BNM-1	298	-0.0332	-245.1	PPB	03 Mar 2021 17:26:13	U	SMPL
M1545-21 - 1	RIM-7.55	363	-0.0171	-245.1	PPB	03 Mar 2021 17:28:30	U	SMPL
M1545-22 - 1	RIM-7.55	406	-0.0066	-245.1	PPB	03 Mar 2021 17:30:46	U	SMPL
M1545-01LX5 - 1		413	-0.0048	-245.1	PPB	03 Mar 2021 17:33:03	U	SMPL
M1545-01A - 1		15984	3.8318	-245.1	PPB	03 Mar 2021 17:35:19	U	SMPL
CCV67 - 1	CCV67	21359	5.1562	-245.1	PPB	03 Mar 2021 17:37:34	U	SMPL
CCB67 - 1	CCB67	329	-0.0255	-245.1	PPB	03 Mar 2021 17:39:49	U	SMPL