



## Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113345

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2718         |
| buffer solution pH 7 yellow    | W2627         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2633         |
| buffer solution pH 7 yellow    | W2719         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.3             | 4.01        | 03/03/2021 | 16:40     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.3             | 7.00        | 03/03/2021 | 16:41     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.3             | 10.03       | 03/03/2021 | 16:42     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.3             | 7.02        | 03/03/2021 | 16:44     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.3             | 2.01        | 03/03/2021 | 16:45     |
| 6   | M1549-01    | 1  | Solid  | 20.03       | 20          | 20.1             | 7.13        | 03/03/2021 | 16:46     |
| 7   | M1549-01DUP | 1  | Solid  | 20.04       | 20          | 20.2             | 7.14        | 03/03/2021 | 16:47     |
| 8   | M1550-01    | 1  | Solid  | 20.05       | 20          | 20.1             | 6.58        | 03/03/2021 | 16:48     |
| 9   | CCV2        | 1  | Water  | NA          | NA          | 20.2             | 12.02       | 03/03/2021 | 16:49     |

✓ 113345

**Date :** 03-03-2021 16:28:37

Date/Time 03/03/2021 17:35  
Received by: CP SM  
Relinquished by: SD/009  
Open

Page 1 of 1

Date/Time 03/03/2021 16:30  
 Received by: SP (of Dean)  
 Relinquished by: SP (of Dean)