



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113349

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/04/2021	09:00
2	CAL2	1	Water	NA	NA	20.2	7.01	03/04/2021	09:01
3	CAL3	1	Water	NA	NA	20.2	10.03	03/04/2021	09:02
4	ICV	1	Water	NA	NA	20.2	6.98	03/04/2021	09:05
5	CCV1	1	Water	NA	NA	20.2	2.01	03/04/2021	09:06
6	M1552-03	1	Water	NA	NA	20.2	7.16	03/04/2021	09:09
7	M1552-03DUP	1	Water	NA	NA	20.3	7.17	03/04/2021	09:10
8	M1552-04	1	Water	NA	NA	20.4	7.01	03/04/2021	09:11
9	M1553-05	1	Water	NA	NA	20.1	7.36	03/04/2021	09:12
10	M1553-06	1	Water	NA	NA	20.1	8.25	03/04/2021	09:15
11	M1553-07	1	Water	NA	NA	20.3	7.94	03/04/2021	09:16
12	M1553-08	1	Water	NA	NA	20.3	9.40	03/04/2021	09:19
13	M1553-09	1	Water	NA	NA	20.4	5.24	03/04/2021	09:20
14	CCV2	1	Water	NA	NA	20.2	12.02	03/04/2021	09:22

WORKLIST(Hardcopy Internal Chain)

VB 113349

WorkList Name : ph m1552 WorkList ID : 146841 Department : Wet-Chemistry Date : 03-04-2021 08:42:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/03/2021	Water	M1552-03	pH	Cool 4 deg C	PSEG03	I61	CHVB0435	03/03/2021	9040C
03/03/2021	Water	M1552-04	pH	Cool 4 deg C	PSEG03	I61	20416	03/03/2021	9040C
03/03/2021	Water	M1553-05	pH	Cool 4 deg C	PSEG03	I63	1730	03/03/2021	9040C
03/03/2021	Water	M1553-06	pH	Cool 4 deg C	PSEG03	I63	1731	03/03/2021	9040C
03/03/2021	Water	M1553-07	pH	Cool 4 deg C	PSEG03	I63	30-31	03/03/2021	9040C
03/03/2021	Water	M1553-08	pH	Cool 4 deg C	PSEG03	I63	1729	03/03/2021	9040C
03/03/2021	Water	M1553-09	pH	Cool 4 deg C	PSEG03	I63	1742	03/03/2021	9040C

Date/Time 03/04/2021 8:45
 Received by: JP (wet Chem)
 Relinquished by: KS (wet Chem)

Date/Time 03/04/2021 10:00
 Received by: KS (wet Chem)
 Relinquished by: JP (wet Chem)