



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113350

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	03/04/2021	09:20
2	M1507-01		1	0	0.00	32.0	32.3	03/04/2021	09:30
3	M1507-01DUP		1	0	0.00	32.0	32.3	03/04/2021	09:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB113350

WorkList Name : FP3221

WorkList ID : 146755

Department : Wet-Chemistry

Date : 03-01-2021 16:02:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2021	Solid	M1507-01	Flash Point	Cool 4 deg C	PSEG03	I33	40959	02/26/2021	1010A

Date/Time 03/04/21 08:55
 Received by: KS (watcher)
 Relinquished by: JN (watcher)

Date/Time 03/04/21 09:45
 Received by: JN (watcher)
 Relinquished by: KS (watcher)