



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113351

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	03/04/2021	09:20
2	M1538-01		1	9	100.00	>212.0	>212.0	03/04/2021	10:05
3	M1538-01DUP		1	9	100.00	>212.0	>212.0	03/04/2021	10:35
4	M1541-02		1	10	100.00	>212.0	>212.0	03/04/2021	10:58
5	M1545-01		1	11	100.00	>212.0	>212.0	03/04/2021	11:20
6	M1545-02		1	10	100.00	>212.0	>212.0	03/04/2021	11:45
7	M1545-03		1	11	100.00	>212.0	>212.0	03/04/2021	12:10
8	M1545-06		1	10	100.00	>212.0	>212.0	03/04/2021	12:35
9	M1545-07		1	11	100.00	>212.0	>212.0	03/04/2021	12:58
10	M1545-08		1	15	100.00	>212.0	>212.0	03/04/2021	13:20
11	M1545-20		1	14	100.00	>212.0	>212.0	03/04/2021	13:50
12	M1545-21		1	15	100.00	>212.0	>212.0	03/04/2021	14:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP3221-1 WorkList ID : 146793 Department : Wet-Chemistry Date : 03-02-2021 14:47:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/09/2021	Water	M1538-01	Flash Point	Cool 4 deg C	PSEG03	I51	1-2-TOTES	03/02/2021	1010A
03/04/2021	Water	M1541-02	Flash Point	Cool 4 deg C	PSEG03	I51	FM-786	03/02/2021	1010A
03/09/2021	Water	M1545-01	Flash Point	Cool 4 deg C	TULL02	I13	MH-5	03/02/2021	1010A
03/09/2021	Water	M1545-02	Flash Point	Cool 4 deg C	TULL02	I13	MH-725	03/02/2021	1010A
03/09/2021	Water	M1545-03	Flash Point	Cool 4 deg C	TULL02	I13	MH-20A	03/02/2021	1010A
03/09/2021	Water	M1545-06	Flash Point	Cool 4 deg C	TULL02	I13	MH-F	03/02/2021	1010A
03/09/2021	Water	M1545-07	Flash Point	Cool 4 deg C	TULL02	I13	MH-J	03/02/2021	1010A
03/09/2021	Water	M1545-08	Flash Point	Cool 4 deg C	TULL02	I13	MH-G	03/02/2021	1010A
03/09/2021	Water	M1545-20	Flash Point	Cool 4 deg C	TULL02	I13	BNM-1	03/02/2021	1010A
03/09/2021	Water	M1545-21	Flash Point	Cool 4 deg C	TULL02	I13	RIM-7.55	03/02/2021	1010A

Date/Time 03/04/21 08:55
 Received by: KS (Wetchem)
 Relinquished by: SP (Wetchem)

Date/Time 03/04/21 14:20
 Received by: SP (Wetchem)
 Relinquished by: KS (Wetchem)