



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113367

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	03/05/2021	09:15
2	M1553-01		1	8	58.00	136.4	136.7	03/05/2021	09:45
3	M1553-01DUP		1	8	59.00	138.2	138.5	03/05/2021	10:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 113367

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp3421 WorkList ID : 146872 Department : Wet-Chemistry Date : 03-04-2021 11:28:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/10/2021	Solid	M1553-01	Flash Point	Cool 4 deg C	PSEG03	I63	1736	03/03/2021	1010A

Date/Time 03/05/21 08:45
Received by: KS (Welch)
Relinquished by: RC (Ext 196)

Date/Time 03/05/21 10:15
Received by: RC (Ext 196)
Relinquished by: KS (Welch)