


LB113370

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.021	0.021	27	0.000	27					-
0.2	0.200	0.211	0.011	806	0.0 %	806					6
2.5	2.500	2.379	-0.121	9735	0.0 %	9735					-5
5.0	5.000	5.053	0.053	20741	0.0 %	20741					1
7.5	7.500	7.652	0.152	31442	0.0 %	31442					2
10.0	10.000	9.884	-0.116	40633	0.0 %	40633					-1

LB113370 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Mar 2021 11:16:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.0000	7470A	PPB	05 Mar 2021 11:19:44	S	Std
0.2 - 1	50.2	806	-	0.2000	7470A	PPB	05 Mar 2021 11:22:00	S	Std
2.5 - 1	52.5	9735	-	2.5000	7470A	PPB	05 Mar 2021 11:24:17	S	Std
5.0 - 1	55	20741	-	5.0000	7470A	PPB	05 Mar 2021 11:26:34	S	Std
7.5 - 1	57.5	31442	-	7.5000	7470A	PPB	05 Mar 2021 11:28:50	S	Std
10.0 - 1	510	40633	-	10.0000	7470A	PPB	05 Mar 2021 11:33:41	S	Std
ICV25 - 1	ICV25	15093	3.6808	-	7470A	PPB	05 Mar 2021 11:36:24	U	SMPL
ICB25 - 1	ICB25	-16	0.0109	-	7470A	PPB	05 Mar 2021 11:38:39	U	SMPL
CCV73 - 1	CCV73	20987	5.1124	-	7470A	PPB	05 Mar 2021 11:40:53	U	SMPL
CCB73 - 1	CCB73	-20	0.0099	-	7470A	PPB	05 Mar 2021 11:43:08	U	SMPL
CRA - 1	CRA	810	0.2115	-	7470A	PPB	05 Mar 2021 11:45:22	U	SMPL
HighStd - 1	HighStd	40937	9.9581	-	7470A	PPB	05 Mar 2021 11:47:37	U	SMPL
ChkStd - 1	ChkStd	29477	7.1746	-	7470A	PPB	05 Mar 2021 11:49:52	U	SMPL
PB134981BL - 1	PBW	-2	0.0143	-	7470A	PPB	05 Mar 2021 11:52:08	U	SMPL
PB134981BS - 1	LCSW	14989	3.6555	-	7470A	PPB	05 Mar 2021 11:54:24	U	SMPL
M1552-05 - 1	435-416	67	0.0310	-	7470A	PPB	05 Mar 2021 11:56:40	U	SMPL
M1552-05DUP - 1	435-416DUP	99	0.0388	-	7470A	PPB	05 Mar 2021 11:58:56	U	SMPL
M1552-05MS - 1	435-416MS	16683	4.0670	-	7470A	PPB	05 Mar 2021 12:01:12	U	SMPL
M1552-05MSD - 1	435-416MSD	16892	4.1177	-	7470A	PPB	05 Mar 2021 12:03:29	U	SMPL
M1553-07 - 1	30-31	22	0.0201	-	7470A	PPB	05 Mar 2021 12:05:45	U	SMPL
CCV74 - 1	CCV74	20985	5.1119	-	7470A	PPB	05 Mar 2021 12:08:00	U	SMPL
CCB74 - 1	CCB74	0	0.0148	-	7470A	PPB	05 Mar 2021 12:10:16	U	SMPL
M1553-08 - 1	1729	46	0.0259	-	7470A	PPB	05 Mar 2021 12:12:31	U	SMPL
M1553-09 - 1	1742	34	0.0230	-	7470A	PPB	05 Mar 2021 12:14:46	U	SMPL
PB134981TB - 1	PB134970TB	90	0.0366	-	7470A	PPB	05 Mar 2021 12:17:01	U	SMPL
M1552-05LX5 - 1		61	0.0296	-	7470A	PPB	05 Mar 2021 12:19:17	U	SMPL
M1552-05A - 1		16350	3.9861	-	7470A	PPB	05 Mar 2021 12:21:33	U	SMPL
CCV75 - 1	CCV75	18636	4.5413	-	7470A	PPB	05 Mar 2021 12:23:49	U	SMPL
CCB75 - 1	CCB75	-276	-0.0523	-	7470A	PPB	05 Mar 2021 12:26:05	U	SMPL