



Analytical Summary Report

Analysis Method: 120.1

ANALYST: Alexander

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB113372

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: pH MeterTempProbe

Reagent/Standard	Lot/Log #
conductivity standard	WP89507

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB113372BL		1	19.8	0.82	0.91	03/05/2021	11:10
2	LB113372BS	1413	1	20.7	1290.00	1405.43	03/05/2021	11:14
3	M1561-08		1	20.0	694.00	767.27	03/05/2021	11:18

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$