



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113375

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/05/2021	12:15
2	CAL2	1	Water	NA	NA	20.3	7.00	03/05/2021	12:16
3	CAL3	1	Water	NA	NA	20.4	10.02	03/05/2021	12:17
4	ICV	1	Water	NA	NA	20.4	6.99	03/05/2021	12:18
5	CCV1	1	Water	NA	NA	20.3	2.01	03/05/2021	12:19
6	M1585-02	1	Water	NA	NA	20.4	7.14	03/05/2021	13:00
7	M1585-02DUP	1	Water	NA	NA	20.5	7.15	03/05/2021	13:01
8	M1586-01	1	Water	NA	NA	20.6	6.46	03/05/2021	13:05
9	M1586-02	1	Water	NA	NA	20.5	5.84	03/05/2021	13:07
10	M1586-03	1	Water	NA	NA	20.5	5.80	03/05/2021	13:10
11	CCV2	1	Water	NA	NA	20.3	12.02	03/05/2021	13:15

WB 113375

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1586 WorkList ID : 146930 Department : Wet-Chemistry Date : 03-05-2021 12:01:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/04/2021	Water	M1585-02 - D	pH	Cool 4 deg C	TETR06	J11	TT-028-IDWSO-20210304	03/04/2021	9040C
03/03/2021	Water	M1586-01 - A	pH	Cool 4 deg C	STER07	J12	CELL-A	03/03/2021	9040C
03/03/2021	Water	M1586-02 - A	pH	Cool 4 deg C	STER07	J12	CELL-B/C	03/03/2021	9040C
03/03/2021	Water	M1586-03 - A	pH	Cool 4 deg C	STER07	J12	CELL-D	03/03/2021	9040C

Date/Time 03/05/2021 12:05
Received by: JD (copy)
Relinquished by: RJ (FET - copy)

Date/Time 03/05/2021 13:30
Received by: RJ (FET - copy)
Relinquished by: JD (copy)