



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113376

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/05/2021	12:15
2	CAL2	1	Water	NA	NA	20.3	7.00	03/05/2021	12:16
3	CAL3	1	Water	NA	NA	20.4	10.02	03/05/2021	12:17
4	ICV	1	Water	NA	NA	20.4	6.99	03/05/2021	12:18
5	CCV1	1	Water	NA	NA	20.3	2.01	03/05/2021	12:19
6	M1575-01	1	Water	NA	NA	20.2	12.08	03/05/2021	12:22
7	M1575-01DUP	1	Water	NA	NA	20.3	12.09	03/05/2021	12:23
8	M1575-02	1	Water	NA	NA	20.1	12.36	03/05/2021	12:25
9	M1575-03	1	Water	NA	NA	20.3	12.03	03/05/2021	12:28
10	M1575-04	1	Water	NA	NA	20.3	11.77	03/05/2021	12:30
11	M1575-05	1	Water	NA	NA	20.3	11.84	03/05/2021	12:32
12	CCV2	1	Water	NA	NA	20.3	12.02	03/05/2021	12:33

WORKLIST(Hardcopy Internal Chain)

JB 113376

WorkList Name : CORROSIONITY M1575 WorkList ID : 146931 Department : Wet-Chemistry Date : 03-05-2021 12:02:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/03/2021	Water	M1575-01	F Corrosivity	Cool 4 deg C	JACO05	J11	IDW-GW803-AQ-030321	03/03/2021	9040C
03/03/2021	Water	M1575-02	F Corrosivity	Cool 4 deg C	JACO05	J11	IDW-GW804-AQ-030321	03/03/2021	9040C
03/03/2021	Water	M1575-03	F Corrosivity	Cool 4 deg C	JACO05	J11	IDW-GW805-AQ-030321	03/03/2021	9040C
03/03/2021	Water	M1575-04	F Corrosivity	Cool 4 deg C	JACO05	J11	IDW-GW806-AQ-030321	03/03/2021	9040C
03/04/2021	Water	M1575-05	F Corrosivity	Cool 4 deg C	JACO05	J11	IDW-GW807B-AQ-030421	03/04/2021	9040C

Date/Time 03/05/2021 12:10
 Received by: JD (leaf) (hpm)
 Relinquished by: RJ (leaf) (hpm)

Date/Time 03/05/2021 12:00
 Received by: RJ (leaf) (hpm)
 Relinquished by: JD (leaf) (hpm)