



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113377

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/05/2021	12:15
2	CAL2	1	Water	NA	NA	20.3	7.00	03/05/2021	12:16
3	CAL3	1	Water	NA	NA	20.4	10.02	03/05/2021	12:17
4	ICV	1	Water	NA	NA	20.4	6.99	03/05/2021	12:18
5	CCV1	1	Water	NA	NA	20.3	2.01	03/05/2021	12:19
6	M1550-01	1	Solid	20.02	20	20.3	6.58	03/05/2021	12:35
7	M1576-01	1	Solid	20.03	20	20.1	9.14	03/05/2021	12:37
8	M1576-02	1	Solid	20.04	20	20.3	12.05	03/05/2021	12:40
9	M1576-03	1	Solid	20.02	20	20.4	10.88	03/05/2021	12:42
10	M1576-04	1	Solid	20.03	20	20.4	9.20	03/05/2021	12:45
11	M1576-05	1	Solid	20.3	20	20.4	9.77	03/05/2021	12:48
12	M1576-05DUP	1	Solid	20.4	20	20.5	9.78	03/05/2021	12:49
13	CCV2	1	Water	NA	NA	20.3	12.02	03/05/2021	12:50

28113371

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M1576 WorkList ID : 146932 Department : Wet-Chemistry Date : 03-05-2021 12:03:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/03/2021	Solid	M1550-01- <u>P</u>	Corrosivity	Cool 4 deg C	PSEG03	I61	VNJ-224	03/03/2021	9045D
03/04/2021	Solid	M1576-01 <u>E</u>	Corrosivity	Cool 4 deg C	JACO05	J11	IDW-GW803-SO-030421	03/04/2021	9045D
03/04/2021	Solid	M1576-02 <u>E</u>	Corrosivity	Cool 4 deg C	JACO05	J11	IDW-GW804-SO-030421	03/04/2021	9045D
03/04/2021	Solid	M1576-03 <u>E</u>	Corrosivity	Cool 4 deg C	JACO05	J11	IDW-GW805-SO-030421	03/04/2021	9045D
03/04/2021	Solid	M1576-04 <u>E</u>	Corrosivity	Cool 4 deg C	JACO05	J11	IDW-GW806-SO-030421	03/04/2021	9045D
03/04/2021	Solid	M1576-05 <u>E</u>	Corrosivity	Cool 4 deg C	JACO05	J11	IDW-GW807B-SO-030421	03/04/2021	9045D

Date/Time 03/05/2021 12:05
 Received by: JR Aoy (JRM)
 Relinquished by: RJ CFT-6060

Date/Time 03/05/2021 12:00
 Received by: RJ CFT-6060
 Relinquished by: JR Aoy
Chem