

LB113380


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53129



A= 0.0000e+000

B= 1.8740e-004

C= 4.3918e-003

Rho= 0.9998933

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%d
0.0	0.000	0.011	0.011	35	0.000	35					
0.2	0.200	0.201	0.001	1051	0.0 %	1051					
2.5	2.500	2.418	-0.082	12881	0.0 %	12881					
5.0	5.000	5.094	0.094	27161	0.0 %	27161					
7.5	7.500	7.514	0.014	40073	0.0 %	40073					
10.0	10.000	9.961	-0.039	53129	0.0 %	53129					

1
3
2
0
0

LB113380 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Mar 2021 13:11:24

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.0000	7471B	PPB	05 Mar 2021 13:41:31	S	Std
0.2 - 1	50.2	1051	-	0.2000	7471B	PPB	05 Mar 2021 13:43:48	S	Std
2.5 - 1	52.5	12881	-	2.5000	7471B	PPB	05 Mar 2021 13:46:05	S	Std
5.0 - 1	55	27161	-	5.0000	7471B	PPB	05 Mar 2021 13:48:21	S	Std
7.5 - 1	57.5	40073	-	7.5000	7471B	PPB	05 Mar 2021 13:50:38	S	Std
10.0 - 1	59	53129	-	10.0000	7471B	PPB	05 Mar 2021 13:52:53	S	Std
ICV26 - 1	ICV26	21443	4.0228	-7471B	PPB	PPB	05 Mar 2021 13:55:39	U	SMPL
ICB26 - 1	ICB26	-36	-0.0024	-7471B	PPB	PPB	05 Mar 2021 13:57:54	U	SMPL
CCV76 - 1	CCV76	27624	5.1812	-7471B	PPB	PPB	05 Mar 2021 14:00:09	U	SMPL
CCB76 - 1	CCB76	-55	-0.0059	-7471B	PPB	PPB	05 Mar 2021 14:02:24	U	SMPL
CRA - 1	CRA	1051	0.2014	-7471B	PPB	PPB	05 Mar 2021 14:04:39	U	SMPL
HighStd - 1	HighStd	53367	10.0054	-7471B	PPB	PPB	05 Mar 2021 14:06:54	U	SMPL
ChkStd - 1	ChkStd	38011	7.1277	-7471B	PPB	PPB	05 Mar 2021 14:09:09	U	SMPL
PB135018BL - 1	PBS	-34	-0.0020	-7471B	PPB	PPB	05 Mar 2021 14:11:26	U	SMPL
PB135018BS - 1	LCSS	21970	4.1216	-7471B	PPB	PPB	05 Mar 2021 14:13:43	U	SMPL
M1458-03 - 1	MDL-SOIL-03-QT1-2021	223	0.0462	-7471B	PPB	PPB	05 Mar 2021 14:15:59	U	SMPL
M1551-01 - 1	VEOLIA-VNJ-204	25334	4.7520	-7471B	PPB	PPB	05 Mar 2021 14:20:46	U	SMPL
M1556-01 - 1	TR-04-030321	1734	0.3293	-7471B	PPB	PPB	05 Mar 2021 14:23:02	U	SMPL
M1556-01DUP - 1	TR-04-030321DUP	1725	0.3277	-7471B	PPB	PPB	05 Mar 2021 14:25:20	U	SMPL
M1556-01MS - 1	TR-04-030321MS	24185	4.5367	-7471B	PPB	PPB	05 Mar 2021 14:27:35	U	SMPL
CCV77 - 1	CCV77	27739	5.2027	-7471B	PPB	PPB	05 Mar 2021 14:29:51	U	SMPL
CCB77 - 1	CCB77	-55	-0.0059	-7471B	PPB	PPB	05 Mar 2021 14:32:06	U	SMPL
M1556-01MSD - 1	TR-04-030321MSD	23679	4.4419	-7471B	PPB	PPB	05 Mar 2021 14:34:21	U	SMPL
M1565-01 - 1	TP-1	570790	HIGH	-7471B	PPB	PPB	05 Mar 2021 14:36:36	X	
M1565-03 - 1	TP-2	24994	4.6883	-7471B	PPB	PPB	05 Mar 2021 14:38:24	U	SMPL
M1565-05 - 1	TP-3	55897	10.4796	-7471B	PPB	PPB	05 Mar 2021 14:40:43	U	SMPL
M1565-07 - 1	TP-4	17981	3.3741	-7471B	PPB	PPB	05 Mar 2021 14:42:58	U	SMPL
M1566-01 - 1	RO-282	1173	0.2242	-7471B	PPB	PPB	05 Mar 2021 14:45:16	U	SMPL
M1566-03 - 1	VEOLIA-72-11977	1086	0.2079	-7471B	PPB	PPB	05 Mar 2021 14:47:32	U	SMPL
M1567-01 - 1	PL-02-030421	1240	0.2368	-7471B	PPB	PPB	05 Mar 2021 14:49:49	U	SMPL
M1568-01 - 1	SOUTHSIDE	169	0.0361	-7471B	PPB	PPB	05 Mar 2021 14:52:06	U	SMPL
M1568-03 - 1	EASTSIDE	113	0.0256	-7471B	PPB	PPB	05 Mar 2021 14:54:21	U	SMPL
CCV78 - 1	CCV78	27655	5.1870	-7471B	PPB	PPB	05 Mar 2021 14:56:37	U	SMPL
CCB78 - 1	CCB78	-40	-0.0031	-7471B	PPB	PPB	05 Mar 2021 14:58:52	U	SMPL
M1568-05 - 1	NORTHSIDE	192	0.0404	-7471B	PPB	PPB	05 Mar 2021 15:01:08	U	SMPL
M1565-01DLX10 - 1		110951	20.7968	-7471B	PPB	PPB	05 Mar 2021 15:03:23	U	SMPL
M1565-05DLX2 - 1		27361	5.1319	-7471B	PPB	PPB	05 Mar 2021 15:05:38	U	SMPL
M1556-01LX5 - 1		281	0.0571	-7471B	PPB	PPB	05 Mar 2021 15:07:56	U	SMPL
M1556-01A - 1		24249	4.5487	-7471B	PPB	PPB	05 Mar 2021 15:10:13	U	SMPL
M1565-01DL2X50 - 1		22843	4.2852	-7471B	PPB	PPB	05 Mar 2021 15:12:29	U	SMPL
CCV79 - 1	CCV79	27541	5.1656	-7471B	PPB	PPB	05 Mar 2021 15:14:45	U	SMPL
CCB79 - 1	CCB79	-53	-0.0055	-7471B	PPB	PPB	05 Mar 2021 15:17:02	U	SMPL