



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113383

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.00	03/05/2021	14:00
2	CAL2	1	Water	NA	NA	20.4	7.01	03/05/2021	14:01
3	CAL3	1	Water	NA	NA	20.4	10.03	03/05/2021	14:02
4	ICV	1	Water	NA	NA	20.4	7.00	03/05/2021	14:05
5	CCV1	1	Water	NA	NA	20.4	2.01	03/05/2021	14:06
6	M1591-01	1	Solid	20.03	20	20.3	8.06	03/05/2021	14:10
7	M1591-01DUP	1	Solid	20.04	20	20.4	8.07	03/05/2021	14:11
8	CCV2	1	Water	NA	NA	20.3	12.03	03/05/2021	14:13