

LB113385

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AM

Instrument ID : Konelab

3/5/2021 14:50

Test: Total CN

Sample Id	^{ug/L} Result	Dil. 1 +	Response	Errors
ICV1	95.298	0.0	0.083	
ICB1	1.782	0.0	0.001	
CCV1	241.923	0.0	0.212	
CCB1	1.404	0.0	0.000	
PB135006BL	1.222	0.0	0.000	
M1550-01	1.269	0.0	0.000	
M1550-01DUP	1.225	0.0	0.000	
CCV2	242.196	0.0	0.212	
CCB2	1.221	0.0	0.000	

N	9
Mean	65.282
SD	104.8348
CV%	160.59

Aquakem v. 7.2AQ1

Results from time period:

Fri Mar 05 14:39:54 2021

Fri Mar 05 14:45:43 2021

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	1.0039	µg/l	3/5/2021 10:45:16	
5.OPPBCN	A	Reactive CN	P	5.8935	µg/l	3/5/2021 10:45:17	
10PPBCN	A	Reactive CN	P	10.3516	µg/l	3/5/2021 10:45:18	
50PPBCN	A	Reactive CN	P	47.9488	µg/l	3/5/2021 10:45:19	
100PPBCN	A	Reactive CN	P	99.652	µg/l	3/5/2021 10:45:20	
250PPBCN	A	Reactive CN	P	249.7831	µg/l	3/5/2021 10:45:21	
500PPBCN	A	Reactive CN	P	500.3672	µg/l	3/5/2021 10:45:22	
ICV1	S	Reactive CN	P	95.2981	µg/l	3/5/2021 14:39:55	
ICB1	S	Reactive CN	P	1.7816	µg/l	3/5/2021 14:39:57	
CCV1	S	Reactive CN	P	241.9235	µg/l	3/5/2021 14:39:58	
CCB1	S	Reactive CN	P	1.4043	µg/l	3/5/2021 14:40:01	
PB135006BL	S	Reactive CN	P	1.2222	µg/l	3/5/2021 14:40:02	
M1550-01	S	Reactive CN	P	1.2694	µg/l	3/5/2021 14:40:04	
M1550-01DUP	S	Reactive CN	P	1.2251	µg/l	3/5/2021 14:45:38	
CCV2	S	Reactive CN	P	242.1959	µg/l	3/5/2021 14:45:41	
CCB2	S	Reactive CN	P	1.2211	µg/l	3/5/2021 14:45:43	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AM Instrument ID : Konelab

3/5/2021 10:48


Test Total CN

Accepted 3/5/2021 10:47

Factor 1138
Bias -0.001

Coeff. of det. 0.999969

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors %D
1	0.0PPBCN	-0.000	1.0039	0.0000	18
2	5.0PPBCN	0.004	5.8935	5.0000	4
3	10PPBCN	0.008	10.3516	10.0000	4
4	50PPBCN	0.041	47.9488	50.0000	10
5	100PPBCN	0.087	99.6520	100.0000	0
6	250PPBCN	0.219	249.7831	250.0000	0
7	500PPBCN	0.439	500.3672	500.0000	0

03/05/2021
AM