

LB 113386


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

47988



A= 0.0000e+000

B= 2.0795e-004 slope

C= -4.6348e-003 y-intercept

Rho= 0.9999660

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.00	0.000	0.012	0.012	81	0.000	81					1
0.05	0.050					332					-2
0.20	0.200	0.196	-0.004	967	0.0 %	967					-2
2.50	2.500	2.452	-0.048	11815	0.0 %	11815					1
5.00	5.000	5.046	0.046	24287	0.0 %	24287					0
7.50	7.500	7.519	0.019	36181	0.0 %	36181					0
10.0	10.000	9.974	-0.026	47988	0.0 %	47988					0

LB113386
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	81	-		0 SFAM01.1	PPB	3/5/2021 15:46	Std
	0.2 S01	967	-		0.2 SFAM01.1	PPB	3/5/2021 15:49	Std
	2.5 S02	11815	-		2.5 SFAM01.1	PPB	3/5/2021 15:51	Std
	5 S03	24287	-		5 SFAM01.1	PPB	3/5/2021 15:53	Std
	7.5 S04	36181	-		7.5 SFAM01.1	PPB	3/5/2021 15:56	Std
	10 S05	47988	-		10 SFAM01.1	PPB	3/5/2021 15:58	Std
ICV018	ICV018	18881	3.9216 -		SFAM01.1	PPB	3/5/2021 16:01	SMPL
ICB018	ICB018	63	0.0085 -		SFAM01.1	PPB	3/5/2021 16:03	SMPL
CCV080	CCV080	24529	5.0961 -		SFAM01.1	PPB	3/5/2021 16:05	SMPL
CCB080	CCB080	59	0.0076 -		SFAM01.1	PPB	3/5/2021 16:07	SMPL
PB135017BL	PBS017	95	0.0151 -		SFAM01.1	PPB	3/5/2021 16:09	SMPL
M1495-19	X3107	423102	HIGH -		SFAM01.1	PPB	3/5/2021 16:12	
M1495-20	X3107D	427208	HIGH -		SFAM01.1	PPB	3/5/2021 16:19	
M1495-21	X3107S	429715	HIGH -		SFAM01.1	PPB	3/5/2021 16:26	
CCV081	CCV081	24234	5.0347 -		SFAM01.1	PPB	3/5/2021 16:28	SMPL
CCB081	CCB081	-19	-0.0086 -		SFAM01.1	PPB	3/5/2021 16:31	SMPL
M1495-19DLX10	X3107	65040	13.5202 -		SFAM01.1	PPB	3/5/2021 16:33	SMPL
M1495-20DLX10	X3107D	64868	13.4844 -		SFAM01.1	PPB	3/5/2021 16:42	SMPL
M1495-21DLX10	X3107S	68445	14.2282 -		SFAM01.1	PPB	3/5/2021 16:51	SMPL
CCV082	CCV082	24825	5.1576 -		SFAM01.1	PPB	3/5/2021 16:53	SMPL
CCB082	CCB082	15	-0.0015 -		SFAM01.1	PPB	3/5/2021 16:56	SMPL
M1495-19DL2X25	X3107	26569	5.5203 -		SFAM01.1	PPB	3/5/2021 16:58	SMPL
M1495-20DL2X25	X3107D	27054	5.6211 -		SFAM01.1	PPB	3/5/2021 17:07	SMPL
M1495-21DL2X25	X3107S	27973	5.8122 -		SFAM01.1	PPB	3/5/2021 17:16	SMPL
CCV083	CCV083	24635	5.1181 -		SFAM01.1	PPB	3/5/2021 17:18	SMPL
CCB083	CCB083	51	0.006 -		SFAM01.1	PPB	3/5/2021 17:20	SMPL