



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113387

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/05/2021	16:15
2	CAL2	1	Water	NA	NA	20.4	7.00	03/05/2021	16:16
3	CAL3	1	Water	NA	NA	20.4	10.03	03/05/2021	16:17
4	ICV	1	Water	NA	NA	20.4	6.98	03/05/2021	16:19
5	CCV1	1	Water	NA	NA	20.5	2.01	03/05/2021	16:20
6	M1578-03	1	Water	NA	NA	20.4	5.62	03/05/2021	16:22
7	M1578-03DUP	1	Water	NA	NA	20.5	5.63	03/05/2021	16:23
8	M1578-04	1	Water	NA	NA	20.3	5.60	03/05/2021	16:25
9	M1578-05	1	Water	NA	NA	20.2	5.58	03/05/2021	16:27
10	M1578-06	1	Water	NA	NA	20.2	6.64	03/05/2021	16:30
11	M1578-07	1	Water	NA	NA	20.4	6.67	03/05/2021	16:32
12	M1578-08	1	Water	NA	NA	20.2	6.92	03/05/2021	16:35
13	CCV2	1	Water	NA	NA	20.3	12.02	03/05/2021	16:37

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1578 WorkList ID : 146939 Department : Wet-Chemistry Date : 03-05-2021 16:07:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2021	Water	M1578-03	pH	Cool 4 deg C	PSEG03	J13	41737	03/05/2021	9040C
03/05/2021	Water	M1578-04	pH	Cool 4 deg C	PSEG03	J13	30753	03/05/2021	9040C
03/05/2021	Water	M1578-05	pH	Cool 4 deg C	PSEG03	J13	30754	03/05/2021	9040C
03/05/2021	Water	M1578-06	pH	Cool 4 deg C	PSEG03	J13	30755	03/05/2021	9040C
03/05/2021	Water	M1578-07	pH	Cool 4 deg C	PSEG03	J13	30756	03/05/2021	9040C
03/05/2021	Water	M1578-08	pH	Cool 4 deg C	PSEG03	J13	30757	03/05/2021	9040C

Date/Time 03/05/21 16:10
 Received by: SP (Wet Chem)
 Relinquished by: KS (Wet Chem)

Date/Time 03/05/21 17:20
 Received by: KS (Wet Chem)
 Relinquished by: SP (Wet Chem)