



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB113395

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP89817
Residual chlorine ICV-LCS, 0.4PPM	WP89816
Chlorine Calibration std, 0.1ppm	WP89811
Chlorine Calibration std, 0.2ppm	WP89812
Chlorine Calibration std, 0.8ppm	WP89814
Chlorine Calibration std, 0.0ppm	WP89808
Chlorine Calibration std, 1.6ppm	WP89815
Residual Chlorine LOQ 0.1ppm	WP89818
Residual Chlorine Calibration and CCV std, 0	WP89813
Total Chlorine Powder Pillows	W2249

Intercept: 0.0103

Slope: 1.0446

Regression: 0.999915

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		03/08/2021	09:50
2	CAL2	0.1	1	0.000	0.120	0.120	0.11	-5	03/08/2021	09:53
3	CAL3	0.2	1	0.000	0.230	0.230	0.21	-5	03/08/2021	09:56
4	CAL4	0.4	1	0.000	0.420	0.420	0.39	2	03/08/2021	09:59
5	CAL5	0.8	1	0.000	0.850	0.850	0.80	-0.5	03/08/2021	10:02
6	CAL6	1.6	1	0.000	1.680	1.680	1.60	0.1	03/08/2021	10:06



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB113395

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4300	0.4300	0.4020	03/08/2021	10:10
2	ICB				1	0.0000	0.0100	0.0100	0.0000	03/08/2021	10:13
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.3920	03/08/2021	10:16
4	CCB1				1	0.0000	0.0100	0.0100	0.0000	03/08/2021	10:20
5	LB113395BL	50	50		1	0.0000	0.0100	0.0100	0.0000	03/08/2021	10:23
6	LB113395BS	50	50	0.4	1	0.0000	0.4400	0.4400	0.4110	03/08/2021	10:27
7	M1458-07	50	50		1	0.0000	0.0600	0.0600	0.0480	03/08/2021	10:30
8	M1458-08	50	50		1	0.0000	0.1200	0.1200	0.1050	03/08/2021	10:33
9	M1561-19	50	50		1	0.0000	2.4000	2.4000	2.2880	03/08/2021	10:36
10	M1561-19DUP	50	50		1	0.0000	2.4100	2.4100	2.2970	03/08/2021	10:39
11	M1561-19				2	0.0000	1.2400	1.2400	1.1770	03/08/2021	10:45
12	M1561-19DUP				2	0.0000	1.2500	1.2500	1.1870	03/08/2021	10:48
13	CCV2			0.4	1	0.0000	0.4300	0.4300	0.4020	03/08/2021	10:52
14	CCB2				1	0.0000	0.0100	0.0100	0.0000	03/08/2021	10:56