

LB113412


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

58856



A= 0.0000e+000

B= 1.6966e-004

C= -3.5764e-003

Rho= 0.9999314

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.005	0.005	52	0.000	52					-
0.2	0.200	0.197	-0.003	1181	0.0 %	1181					-2
2.5	2.500	2.446	-0.054	14440	0.0 %	14440					-2
5.0	5.000	5.087	0.087	30007	0.0 %	30007					2
7.5	7.500	7.482	-0.018	44121	0.0 %	44121					0
10.0	10.000	9.982	-0.018	58856	0.0 %	58856					0

LB113412 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Mar 2021 14:51:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	52	-	0.00007471B	PPB		09 Mar 2021 15:00:25	S	Std
0.2 - 1	50.2	1181	-	0.20007471B	PPB		09 Mar 2021 15:02:42	S	Std
2.5 - 1	52.5	14440	-	2.50007471B	PPB		09 Mar 2021 15:04:59	S	Std
5.0 - 1	55	30007	-	5.00007471B	PPB		09 Mar 2021 15:07:16	S	Std
7.5 - 1	57.5	44121	-	7.50007471B	PPB		09 Mar 2021 15:09:32	S	Std
10.0 - 1	510	58856	-	10.00007471B	PPB		09 Mar 2021 15:11:48	S	Std
ICV29 - 1	ICV29	22396	3.7962	-7471B	PPB		09 Mar 2021 15:14:35	U	SMPL
ICB29 - 1	ICB29	-38	-0.0100	-7471B	PPB		09 Mar 2021 15:16:49	U	SMPL
CCV92 - 1	CCV92	29932	5.0747	-7471B	PPB		09 Mar 2021 15:19:04	U	SMPL
CCB92 - 1	CCB92	-30	-0.0087	-7471B	PPB		09 Mar 2021 15:21:19	U	SMPL
CRA - 1	CRA	1150	0.1915	-7471B	PPB		09 Mar 2021 15:23:34	U	SMPL
HighStd - 1	HighStd	58707	9.9568	-7471B	PPB		09 Mar 2021 15:25:49	U	SMPL
ChkStd - 1	ChkStd	41647	7.0623	-7471B	PPB		09 Mar 2021 15:28:04	U	SMPL
PB135056BL - 1	PBS	-52	-0.0124	-7471B	PPB		09 Mar 2021 15:30:20	U	SMPL
PB135056BS - 1	LCSS	21395	3.6263	-7471B	PPB		09 Mar 2021 15:32:36	U	SMPL
M1577-01 - 1	72-11939-VNJ-229	2630	0.4426	-7471B	PPB		09 Mar 2021 15:34:52	U	SMPL
M1579-01 - 1	HR-02-030521	2828	0.4762	-7471B	PPB		09 Mar 2021 15:37:09	U	SMPL
M1582-01 - 1	VEOLIA-VNJ-208	1171	0.1951	-7471B	PPB		09 Mar 2021 15:39:26	U	SMPL
M1582-03 - 1	RO-363	1024	0.1702	-7471B	PPB		09 Mar 2021 15:41:43	U	SMPL
M1582-07 - 1	VEOLIA-72-11991	1421	0.2375	-7471B	PPB		09 Mar 2021 15:43:58	U	SMPL
CCV93 - 1	CCV93	29736	5.0415	-7471B	PPB		09 Mar 2021 15:46:13	U	SMPL
CCB93 - 1	CCB93	-43	-0.0109	-7471B	PPB		09 Mar 2021 15:48:28	U	SMPL
M1589-01 - 1	FK-GF-02-067-A	7527	1.2735	-7471B	PPB		09 Mar 2021 15:50:44	U	SMPL
M1589-03 - 1	FK-GF-02-067-B	6978	1.1803	-7471B	PPB		09 Mar 2021 15:52:58	U	SMPL
M1590-01 - 1	FK-EF-01-018-ABCD	4839	0.8174	-7471B	PPB		09 Mar 2021 15:55:13	U	SMPL
M1591-01 - 1	FK-PS-01-070	3562	0.6008	-7471B	PPB		09 Mar 2021 15:57:29	U	SMPL
M1595-01 - 1	OF-3-20210305	14235	2.4116	-7471B	PPB		09 Mar 2021 16:04:39	U	SMPL
M1595-01DUP - 1	OF-3-20210305DUP	14154	2.3978	-7471B	PPB		09 Mar 2021 16:06:55	U	SMPL
M1595-02 - 1	M1595-01MS	35892	6.0859	-7471B	PPB		09 Mar 2021 16:09:11	U	SMPL
M1595-03 - 1	M1595-01MSD	36201	6.1384	-7471B	PPB		09 Mar 2021 16:11:28	U	SMPL
M1595-04 - 1	OF-4-20210305	7804	1.3205	-7471B	PPB		09 Mar 2021 16:13:45	U	SMPL
M1595-05 - 1	DUP03-20210305	16385	2.7763	-7471B	PPB		09 Mar 2021 16:19:16	U	SMPL
CCV94 - 1	CCV94	29882	5.0663	-7471B	PPB		09 Mar 2021 16:21:31	U	SMPL
CCB94 - 1	CCB94	-45	-0.0112	-7471B	PPB		09 Mar 2021 16:23:46	U	SMPL
M1595-01LX5 - 1		2860	0.4817	-7471B	PPB		09 Mar 2021 16:26:02	U	SMPL
M1595-01A - 1		34933	5.9232	-7471B	PPB		09 Mar 2021 16:28:17	U	SMPL
CCV95 - 1	CCV95	30080	5.0999	-7471B	PPB		09 Mar 2021 16:30:33	U	SMPL
CCB95 - 1	CCB95	-48	-0.0117	-7471B	PPB		09 Mar 2021 16:32:48	U	SMPL