



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113422

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.0	4.01	03/10/2021	08:25
2	CAL2	1	Water	NA	NA	21.0	7.00	03/10/2021	08:26
3	CAL3	1	Water	NA	NA	21.0	10.03	03/10/2021	08:27
4	ICV	1	Water	NA	NA	21.1	6.99	03/10/2021	08:30
5	CCV1	1	Water	NA	NA	21.1	2.01	03/10/2021	08:31
6	M1561-21	1	Water	NA	NA	22.1	9.45	03/10/2021	08:35
7	M1605-01	1	Water	NA	NA	21.3	5.75	03/10/2021	08:36
8	M1605-02	1	Water	NA	NA	21.3	5.76	03/10/2021	08:37
9	M1605-03	1	Water	NA	NA	21.5	6.53	03/10/2021	08:40
10	M1605-04	1	Water	NA	NA	21.6	6.16	03/10/2021	08:41
11	M1605-05	1	Water	NA	NA	21.5	6.54	03/10/2021	08:43
12	M1605-06	1	Water	NA	NA	21.6	5.55	03/10/2021	08:45
13	M1605-07	1	Water	NA	NA	21.6	6.63	03/10/2021	08:48
14	M1605-08	1	Water	NA	NA	21.5	6.92	03/10/2021	08:49
15	M1605-10	1	Water	NA	NA	21.3	6.17	03/10/2021	08:50
16	CCV2	1	Water	NA	NA	21.4	12.02	03/10/2021	08:51
17	M1605-12	1	Water	NA	NA	20.4	5.70	03/10/2021	08:53
18	M1605-14	1	Water	NA	NA	21.5	5.75	03/10/2021	08:58
19	M1606-01	1	Water	NA	NA	21.3	6.65	03/10/2021	09:00
20	M1606-01DUP	1	Water	NA	NA	21.4	6.66	03/10/2021	09:01

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	21.4	2.01	03/10/2021	09:05

WORKLIST(Hardcopy Internal Chain)

✓ 113422

WorkList Name : PH M1605

WorkList ID : 146995

Department : Wet-Chemistry

Date : 03-10-2021 08:20:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/09/2021	Water	M1605-01	pH	Cool 4 deg C	PSEG03	H22	33013	03/09/2021	9040C
03/09/2021	Water	M1605-02	pH	Cool 4 deg C	PSEG03	H22	33014	03/09/2021	9040C
03/09/2021	Water	M1605-03	pH	Cool 4 deg C	PSEG03	H22	32921	03/09/2021	9040C
03/09/2021	Water	M1605-04	pH	Cool 4 deg C	PSEG03	H22	32926	03/09/2021	9040C
03/09/2021	Water	M1605-05	pH	Cool 4 deg C	PSEG03	H22	32922	03/09/2021	9040C
03/09/2021	Water	M1605-06	pH	Cool 4 deg C	PSEG03	H22	32975	03/09/2021	9040C
03/09/2021	Water	M1605-07	pH	Cool 4 deg C	PSEG03	H22	32978	03/09/2021	9040C
03/09/2021	Water	M1605-08	pH	Cool 4 deg C	PSEG03	H22	32920	03/09/2021	9040C
03/09/2021	Water	M1605-10	pH	Cool 4 deg C	PSEG03	H22	33010	03/09/2021	9040C
03/09/2021	Water	M1605-12	pH	Cool 4 deg C	PSEG03	H22	13-14	03/09/2021	9040C
03/09/2021	Water	M1605-14	pH	Cool 4 deg C	PSEG03	H22	921-978	03/09/2021	9040C
03/09/2021	Water	M1606-01	pH	Cool 4 deg C	PSEG03	H23	OW-SEPARATOR-PIT	03/09/2021	9040C

Date/Time 03/10/2021 8:22
 Received by: SP (wet Chem)
 Relinquished by: CP SMI

Date/Time 03/10/2021 10:00
 Received by: CP SMI
 Relinquished by: SP (wet Chem)