



Analytical Summary Report

Analysis Method: 9041A

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113424

Slope : N/A

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616
3.0-5.5 pH paper refill	W1934
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.0	4.0	03/10/2021	08:25
2	CAL2	1	Water	NA	NA	21.0	7.0	03/10/2021	08:26
3	CAL3	1	Water	NA	NA	21.0	10.0	03/10/2021	08:27
4	ICV	1	Water	NA	NA	21.1	7.0	03/10/2021	08:30
5	CCV1	1	Water	NA	NA	21.1	2.0	03/10/2021	08:31
6	M1561-22	1	Water	NA	NA	22.1	9.4	03/10/2021	08:35
7	CCV2	1	Water	NA	NA	21.4	12.0	03/10/2021	08:51