



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB113426

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP89866
Residual chlorine ICV-LCS, 0.4PPM	WP89865
Chlorine Calibration std, 0.1ppm	WP89862
Chlorine Calibration std, 0.2ppm	WP89863
Chlorine Calibration std, 0.8ppm	WP89867
Chlorine Calibration std, 0.0ppm	WP89861
Chlorine Calibration std, 1.6ppm	WP89868
Residual Chlorine Calibration and CCV std, 0	WP89864
Total Chlorine Powder Pillows	W2249

Intercept: 0.0131

Slope: 1.0584

Regression: 0.999862

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.010	0.010	0.00		03/10/2021	09:30
2	CAL2	0.1	1	0.000	0.120	0.120	0.10	-1	03/10/2021	09:33
3	CAL3	0.2	1	0.000	0.210	0.210	0.19	7	03/10/2021	09:36
4	CAL4	0.4	1	0.000	0.450	0.450	0.41	-3.2	03/10/2021	09:39
5	CAL5	0.8	1	0.000	0.870	0.870	0.81	-1.3	03/10/2021	09:42
6	CAL6	1.6	1	0.000	1.700	1.700	1.59	0.4	03/10/2021	09:46



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB113426

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4400	0.4400	0.4030	03/10/2021	09:50
2	ICB				1	0.0000	0.0100	0.0100	-0.0030	03/10/2021	09:53
3	CCV1			0.4	1	0.0000	0.4400	0.4400	0.4030	03/10/2021	09:56
4	CCB1				1	0.0000	0.0100	0.0100	-0.0030	03/10/2021	10:00
5	LB113426BL	50	50		1	0.0000	0.0200	0.0200	0.0070	03/10/2021	10:03
6	LB113426BS	50	50	0.4	1	0.0000	0.4500	0.4500	0.4130	03/10/2021	10:07
7	M1458-09	50	50		1	0.0000	0.0700	0.0700	0.0540	03/10/2021	10:10
8	CCV2			0.4	1	0.0000	0.4500	0.4500	0.4130	03/10/2021	10:14
9	CCB2				1	0.0000	0.0100	0.0100	-0.0030	03/10/2021	10:18