



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113444

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/10/2021	09:45
2	M1605-08		1	10	100.00	>212.0	>212.0	03/10/2021	10:15
3	M1605-08DUP		1	10	100.00	>212.0	>212.0	03/10/2021	10:40
4	M1605-10		1	9	100.00	>212.0	>212.0	03/10/2021	11:10
5	M1605-12		1	11	100.00	>212.0	>212.0	03/10/2021	11:35
6	M1605-14		1	12	100.00	>212.0	>212.0	03/10/2021	12:05
7	M1606-01		1	11	100.00	>212.0	>212.0	03/10/2021	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 113444

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP31021

WorkList ID : 147027

Department : Wet-Chemistry

Date : 03-10-2021 09:15:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2021	Water	M1605-08	Flash Point	Cool 4 deg C	PSEG03	H22	32920	03/09/2021	1010A
03/12/2021	Water	M1605-10	Flash Point	Cool 4 deg C	PSEG03	H22	33010	03/09/2021	1010A
03/12/2021	Water	M1605-12	Flash Point	Cool 4 deg C	PSEG03	H22	13-14	03/09/2021	1010A
03/12/2021	Water	M1605-14	Flash Point	Cool 4 deg C	PSEG03	H22	921-978	03/09/2021	1010A
03/12/2021	Water	M1606-01	Flash Point	Cool 4 deg C	PSEG03	H23	OW-SEPARATOR-PIT	03/09/2021	1010A

Date/Time 03/10/21 09:25
Received by: KS (wetcher)
Relinquished by: KS (wetcher)

Date/Time 03/10/21 12:40
Received by: KS (wetcher)
Relinquished by: KS (wetcher)