



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113450

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	03/10/2021	16:15
2	CAL2	1	Water	NA	NA	21.1	7.00	03/10/2021	16:16
3	CAL3	1	Water	NA	NA	21.1	10.03	03/10/2021	16:17
4	ICV	1	Water	NA	NA	21.1	6.99	03/10/2021	16:19
5	CCV1	1	Water	NA	NA	21.1	2.01	03/10/2021	16:20
6	M1619-04	1	Water	NA	NA	20.1	6.74	03/10/2021	16:40
7	M1619-04DUP	1	Water	NA	NA	20.2	6.75	03/10/2021	16:41
8	M1619-05	1	Water	NA	NA	20.2	6.75	03/10/2021	16:42
9	M1619-06	1	Water	NA	NA	20.2	6.71	03/10/2021	16:44
10	M1619-07	1	Water	NA	NA	20.1	7.25	03/10/2021	16:45
11	M1619-08	1	Water	NA	NA	20.3	9.05	03/10/2021	16:47
12	CCV2	1	Water	NA	NA	20.2	12.02	03/10/2021	16:48

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1619

WorkList ID : 147039

Department : Wet-Chemistry

Date : 03-10-2021 16:31:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/10/2021	Water	M1619-04	pH	Cool 4 deg C	PSEG03	H33	BUR-1006	03/10/2021	9040C
03/10/2021	Water	M1619-05	pH	Cool 4 deg C	PSEG03	H33	BUR-1007	03/10/2021	9040C
03/10/2021	Water	M1619-06	pH	Cool 4 deg C	PSEG03	H33	BUR-1008	03/10/2021	9040C
03/10/2021	Water	M1619-07	pH	Cool 4 deg C	PSEG03	H33	BUR-1010	03/10/2021	9040C
03/10/2021	Water	M1619-08	pH	Cool 4 deg C	PSEG03	H33	30624	03/10/2021	9040C

Date/Time

Received by:

Relinquished by:

03/10/21 16:35

JP (wet Chem)

KS (wet Chem)

Date/Time

Received by:

Relinquished by:

03/10/21 17:25

KS (wet Chem)

JP (wet Chem)