



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113457

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	03/11/2021	11:20
2	M1619-02		1	10	100.00	>212.0	>212.0	03/11/2021	11:45
3	M1619-02DUP		1	10	100.00	>212.0	>212.0	03/11/2021	12:10
4	M1619-08		1	11	100.00	>212.0	>212.0	03/11/2021	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 113457

WorkList Name : fp31121 WorkList ID : 147060 Department : Wet-Chemistry Date : 03-11-2021 08:58:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/23/2021	Water	M1619-02	Flash Point	Cool 4 deg C	PSEG03	H33	COMP-BUR-1006-BUR-1008,	03/10/2021	1010A
03/23/2021	Water	M1619-08	Flash Point	Cool 4 deg C	PSEG03	H33	30624	03/10/2021	1010A

Date/Time 03/11/21 09:10
 Received by: KS (wetchem)
 Relinquished by: RI (EST-1046)

Date/Time 03/11/21 12:45
 Received by: RI (EST-1046)
 Relinquished by: KS (wetchem)