



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113458

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/11/2021	11:30
2	CAL2	1	Water	NA	NA	20.1	7.00	03/11/2021	11:31
3	CAL3	1	Water	NA	NA	20.1	10.03	03/11/2021	11:32
4	ICV	1	Water	NA	NA	20.2	7.01	03/11/2021	11:33
5	CCV1	1	Water	NA	NA	20.2	2.01	03/11/2021	11:34
6	M1653-01	1	Water	NA	NA	20.3	4.98	03/11/2021	11:35
7	M1653-01DUP	1	Water	NA	NA	20.4	4.99	03/11/2021	11:36
8	CCV2	1	Water	NA	NA	20.3	12.02	03/11/2021	11:37

28 113458

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1563 WorkList ID : 147069 Department : Wet-Chemistry Date : 03-11-2021 11:23:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/06/2021	Water	M1653-01	pH	Cool 4 deg C	STER15	J13	GILROY-OMEGA-CELL	03/06/2021	9040C

Date/Time 03/11/2021 11:25
Received by: SO (wet Chem)
Relinquished by: CF (SM)

Date/Time 03/11/2021 12:25
Received by: CF (SM)
Relinquished by: SO (wet Chem)