



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113459

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	3.99	03/11/2021	11:45
2	CAL2	1	Water	NA	NA	20.3	7.00	03/11/2021	11:46
3	CAL3	1	Water	NA	NA	20.3	10.03	03/11/2021	11:47
4	ICV	1	Water	NA	NA	20.4	7.01	03/11/2021	11:48
5	CCV1	1	Water	NA	NA	20.3	2.01	03/11/2021	11:50
6	M1639-01	1	Solid	20.03	20	20.8	8.31	03/11/2021	12:00
7	M1639-01DUP	1	Solid	20.04	20	20.9	8.31	03/11/2021	12:01
8	CCV2	1	Water	NA	NA	20.2	12.02	03/11/2021	12:03

LB 113459

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1639

WorkList ID : 147074

Department : Wet-Chemistry

Date : 03-11-2021 11:35:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/10/2021	Solid	M1639-01	pH	Cool 4 deg C	TULL01	J42	STOCKPILE	03/10/2021	9045D

Date/Time 03/11/2021 11:40
Received by: JO (ing) (emp)
Relinquished by: JO (ing) (emp)

Date/Time 03/11/2021 12:45
Received by: JO (ing) (emp)
Relinquished by: JO (ing) (emp)