

LB113465


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39097



A= 0.0000e+000

B= 2.5383e-004

C= -3.9514e-002

Rho= 0.9997100

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	0.008	0.008	188	0.000	188					-
0.2	0.200	0.166	-0.034	809	0.0 %	809					-17
2.5	2.500	2.419	-0.081	9687	0.0 %	9687					-3
5.0	5.000	5.140	0.140	20405	0.0 %	20405					3
7.5	7.500	7.582	0.082	30026	0.0 %	30026					1
10.0	10.000	9.885	-0.115	39097	0.0 %	39097					-1

LB113465 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Mar 2021 12:51:57

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	SD	188	-	0.0000	7470A	PPB	11 Mar 2021 13:20:16	S	Std
0.2 - 1	SD.2	809	-	0.2000	7470A	PPB	11 Mar 2021 13:22:33	S	Std
2.5 - 1	SD.5	9687	-	2.5000	7470A	PPB	11 Mar 2021 13:24:50	S	Std
5.0 - 1	SD	20405	-	5.0000	7470A	PPB	11 Mar 2021 13:27:07	S	Std
7.5 - 1	SD.5	30026	-	7.5000	7470A	PPB	11 Mar 2021 13:34:12	S	Std
10.0 - 1	SD	39097	-	10.0000	7470A	PPB	11 Mar 2021 13:36:27	S	Std
ICV33 - 1	ICV33	15319	3.8489	-	7470A	PPB	11 Mar 2021 13:39:15	U	SMPL
ICB33 - 1	ICB33	289	0.0338	-	7470A	PPB	11 Mar 2021 13:41:30	U	SMPL
CCV05 - 1	CCV05	20293	5.1115	-	7470A	PPB	11 Mar 2021 13:43:45	U	SMPL
CCB05 - 1	CCB05	306	0.0382	-	7470A	PPB	11 Mar 2021 13:46:00	U	SMPL
CRA - 1	CRA	784	0.1595	-	7470A	PPB	11 Mar 2021 13:48:14	U	SMPL
HighStd - 1	HighStd	38348	9.6945	-	7470A	PPB	11 Mar 2021 13:52:59	U	SMPL
ChkStd - 1	ChkStd	28024	7.0739	-	7470A	PPB	11 Mar 2021 13:55:14	U	SMPL
PB135117BL - 1	PBW	47	-0.0276	-	7470A	PPB	11 Mar 2021 13:59:58	U	SMPL
PB135117BS - 1	LCSW	15960	4.0117	-	7470A	PPB	11 Mar 2021 14:02:13	U	SMPL
M1560-07 - 1	PT-HG-WP	69586	17.6237	-	7470A	PPB	11 Mar 2021 14:04:30	U	SMPL
M1595-14 - 1	FB02-20210305	-141	-0.0753	-	7470A	PPB	11 Mar 2021 14:06:46	U	SMPL
M1605-12 - 1	13-14	658	0.1275	-	7470A	PPB	11 Mar 2021 14:13:51	U	SMPL
M1605-12DUP - 1	13-14DUP	559	0.1024	-	7470A	PPB	11 Mar 2021 14:16:07	U	SMPL
M1605-12MS - 1	13-14MS	15656	3.9345	-	7470A	PPB	11 Mar 2021 14:18:23	U	SMPL
CCV06 - 1	CCV06	19699	4.9607	-	7470A	PPB	11 Mar 2021 14:20:38	U	SMPL
CCB06 - 1	CCB06	180	0.0062	-	7470A	PPB	11 Mar 2021 14:22:55	U	SMPL
M1605-12MSD - 1	13-14MSD	15991	4.0195	-	7470A	PPB	11 Mar 2021 14:25:10	U	SMPL
M1605-14 - 1	921-978	513	0.0907	-	7470A	PPB	11 Mar 2021 14:27:26	U	SMPL
M1606-01 - 1	OW-SEPARATOR-PIT	773	0.1567	-	7470A	PPB	11 Mar 2021 14:29:42	U	SMPL
M1605-12LX5 - 1		35	-0.0306	-	7470A	PPB	11 Mar 2021 14:41:35	U	SMPL
M1605-12A - 1		16129	4.0546	-	7470A	PPB	11 Mar 2021 14:43:50	U	SMPL
CCV07 - 1	CCV07	19770	4.9788	-	7470A	PPB	11 Mar 2021 14:46:08	U	SMPL
CCB07 - 1	CCB07	222	0.0168	-	7470A	PPB	11 Mar 2021 14:48:24	U	SMPL
M1560-07DLX5 - 1	PT-HG-WP	14676	3.6857	-	7470A	PPB	11 Mar 2021 14:52:55	U	SMPL
CCV08 - 1	CCV08	20228	5.0950	-	7470A	PPB	11 Mar 2021 14:59:41	U	SMPL
CCB08 - 1	CCB08	320	0.0417	-	7470A	PPB	11 Mar 2021 15:01:57	U	SMPL