

LB113469


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38367



A= 0.0000e+000

B= 2.6202e-004

C= -2.3612e-002

Rho= 0.9997864

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.003	-0.003	80	0.000	80					
0.05	0.050	0.044	-0.006	258	0.0 %	258	0				-12
0.2	0.200	0.187	-0.013	803	0.0 %	803					-7
2.5	2.500	2.482	-0.018	9563	0.0 %	9563					-1
5.0	5.000	5.146	0.146	19730	0.0 %	19730					3
7.5	7.500	7.364	-0.136	28195	0.0 %	28195					-2
10.0	10.000	10.029	0.029	38367	0.0 %	38367					0

LB113469 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Mar 2021 15:15:31

Sample ID	Extended ID	μ Abs	Conc	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50.2	80	-	0.0000	245.1	PPB	11 Mar 2021 15:17:57	S	Std
0.05 - 1	50.2	258	-	0.0500	245.1	PPB	11 Mar 2021 15:20:14	S	Std
0.2 - 1	52.5	803	-	0.2000	245.1	PPB	11 Mar 2021 15:22:31	S	Std
2.5 - 1	55	9563	-	2.5000	245.1	PPB	11 Mar 2021 15:24:48	S	Std
5.0 - 1	53.5	19730	-	5.0000	245.1	PPB	11 Mar 2021 15:27:04	S	Std
7.5 - 1	510	28195	-	7.5000	245.1	PPB	11 Mar 2021 15:29:19	S	Std
10.0 - 1		38367	-	10.0000	245.1	PPB	11 Mar 2021 15:47:47	S	Std
ICV34 - 1	ICV34	14754	3.8423	-245.1		PPB	11 Mar 2021 15:50:34	U	SMPL
ICB34 - 1	ICB34	-18	-0.0283	-245.1		PPB	11 Mar 2021 15:52:49	U	SMPL
CCV09 - 1	CCV09	19789	5.1616	-245.1		PPB	11 Mar 2021 15:55:03	U	SMPL
CCB09 - 1	CCB09	-29	-0.0312	-245.1		PPB	11 Mar 2021 15:57:18	U	SMPL
CRA - 1	CRA	257	0.0437	-245.1		PPB	11 Mar 2021 16:07:02	U	SMPL
HighStd - 1	HighStd	39075	10.2150	-245.1		PPB	11 Mar 2021 16:09:17	U	SMPL
ChkStd - 1	ChkStd	26598	6.9457	-245.1		PPB	11 Mar 2021 16:11:32	U	SMPL
PB135118BL - 1	PBW	43	-0.0123	-245.1		PPB	11 Mar 2021 16:16:20	U	SMPL
PB135118BS - 1	LCSW	16343	4.2586	-245.1		PPB	11 Mar 2021 16:18:35	U	SMPL
M1458-10 - 1	LOD-MDL-WATER-04-QT1-2021	648	0.1462	-245.1		PPB	11 Mar 2021 16:20:52	U	SMPL
M1458-12 - 1	MDL-WATER-06-QT1-2021	253	0.0427	-245.1		PPB	11 Mar 2021 16:28:20	U	SMPL
M1559-08 - 1	PT-HG-WP	71357	18.6736	-245.1		PPB	11 Mar 2021 16:30:37	U	SMPL
M1562-02 - 1	COMP	1030	0.2463	-245.1		PPB	11 Mar 2021 16:35:31	U	SMPL
M1562-02DUP - 1	COMPDUP	1046	0.2505	-245.1		PPB	11 Mar 2021 16:37:46	U	SMPL
CCV10 - 1	CCV10	19904	5.1917	-245.1		PPB	11 Mar 2021 16:40:00	U	SMPL
CCB10 - 1	CCB10	-16	-0.0278	-245.1		PPB	11 Mar 2021 16:42:15	U	SMPL
M1562-02MS - 1	COMPMS	13553	3.5276	-245.1		PPB	11 Mar 2021 16:44:31	U	SMPL
M1562-02MSD - 1	COMPMSD	15287	3.9819	-245.1		PPB	11 Mar 2021 16:46:46	U	SMPL
M1562-02LX5 - 1		223	0.0348	-245.1		PPB	11 Mar 2021 16:59:28	U	SMPL
M1562-02A - 1		17261	4.4992	-245.1		PPB	11 Mar 2021 17:01:44	U	SMPL
M1559-08DLX5 - 1	PT-HG-WP	14206	3.6987	-245.1		PPB	11 Mar 2021 17:04:01	U	SMPL
CCV11 - 1	CCV11	20056	5.2315	-245.1		PPB	11 Mar 2021 17:13:03	U	SMPL
CCB11 - 1	CCB11	-5	-0.0249	-245.1		PPB	11 Mar 2021 17:15:19	U	SMPL