



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113472

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/11/2021	15:50
2	CAL2	1	Water	NA	NA	20.2	7.00	03/11/2021	15:51
3	CAL3	1	Water	NA	NA	20.2	10.02	03/11/2021	15:52
4	ICV	1	Water	NA	NA	20.2	6.99	03/11/2021	16:00
5	CCV1	1	Water	NA	NA	20.2	2.01	03/11/2021	16:01
6	M1647-01	1	Water	NA	NA	20.5	6.88	03/11/2021	16:02
7	M1647-01DUP	1	Water	NA	NA	20.6	6.89	03/11/2021	16:03
8	M1647-02	1	Water	NA	NA	20.2	6.10	03/11/2021	16:05
9	M1647-03	1	Water	NA	NA	20.3	6.45	03/11/2021	16:06
10	M1647-04	1	Water	NA	NA	20.1	6.37	03/11/2021	16:07
11	M1647-05	1	Water	NA	NA	20.3	6.83	03/11/2021	16:09
12	M1647-06	1	Water	NA	NA	20.3	5.30	03/11/2021	16:10
13	M1647-07	1	Water	NA	NA	20.4	5.69	03/11/2021	16:11
14	M1647-08	1	Water	NA	NA	20.6	6.40	03/11/2021	16:12
15	M1647-09	1	Water	NA	NA	20.7	6.34	03/11/2021	16:15
16	CCV2	1	Water	NA	NA	20.4	12.02	03/11/2021	16:16
17	M1647-10	1	Water	NA	NA	20.2	6.45	03/11/2021	16:19
18	M1647-11	1	Water	NA	NA	20.4	6.27	03/11/2021	16:20
19	M1647-12	1	Water	NA	NA	20.7	6.12	03/11/2021	16:22
20	M1647-13	1	Water	NA	NA	20.8	5.97	03/11/2021	16:23

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M1647-14	1	Water	NA	NA	20.3	7.02	03/11/2021	16:25
22	M1647-15	1	Water	NA	NA	20.5	6.04	03/11/2021	16:26
23	M1647-16	1	Water	NA	NA	20.4	9.71	03/11/2021	16:27
24	M1648-01	1	Water	NA	NA	20.4	6.57	03/11/2021	16:30
25	M1648-02	1	Water	NA	NA	20.2	6.66	03/11/2021	16:31
26	M1648-03	1	Water	NA	NA	20.3	6.46	03/11/2021	16:33
27	CCV3	1	Water	NA	NA	20.3	2.01	03/11/2021	16:35
28	M1648-04	1	Water	NA	NA	20.6	6.41	03/11/2021	16:36
29	M1648-05	1	Water	NA	NA	20.6	6.44	03/11/2021	16:38
30	M1648-06	1	Water	NA	NA	20.7	6.46	03/11/2021	16:39
31	M1648-06DUP	1	Water	NA	NA	20.8	6.47	03/11/2021	16:40
32	CCV4	1	Water	NA	NA	20.3	12.03	03/11/2021	16:42

WORKLIST(Hardcopy Internal Chain)

LB 113472

WorkList Name : ph m1647

WorkList ID : 147079

Department : Wet-Chemistry

Date : 03-11-2021 15:36:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/11/2021	Water	M1647-01	pH	Cool 4 deg C	PSEG03	H51	41730	03/11/2021	9040C
03/11/2021	Water	M1647-02	pH	Cool 4 deg C	PSEG03	H51	41731	03/11/2021	9040C
03/11/2021	Water	M1647-03	pH	Cool 4 deg C	PSEG03	H51	41732	03/11/2021	9040C
03/11/2021	Water	M1647-04	pH	Cool 4 deg C	PSEG03	H51	41733	03/11/2021	9040C
03/11/2021	Water	M1647-05	pH	Cool 4 deg C	PSEG03	H51	41734	03/11/2021	9040C
03/11/2021	Water	M1647-06	pH	Cool 4 deg C	PSEG03	H51	30788	03/11/2021	9040C
03/11/2021	Water	M1647-07	pH	Cool 4 deg C	PSEG03	H51	30789	03/11/2021	9040C
03/11/2021	Water	M1647-08	pH	Cool 4 deg C	PSEG03	H51	41735	03/11/2021	9040C
03/11/2021	Water	M1647-09	pH	Cool 4 deg C	PSEG03	H51	30780	03/11/2021	9040C
03/11/2021	Water	M1647-10	pH	Cool 4 deg C	PSEG03	H51	30811	03/11/2021	9040C
03/11/2021	Water	M1647-11	pH	Cool 4 deg C	PSEG03	H51	30812	03/11/2021	9040C
03/11/2021	Water	M1647-12	pH	Cool 4 deg C	PSEG03	H51	30813	03/11/2021	9040C
03/11/2021	Water	M1647-13	pH	Cool 4 deg C	PSEG03	H51	30814	03/11/2021	9040C
03/11/2021	Water	M1647-14	pH	Cool 4 deg C	PSEG03	H51	30815	03/11/2021	9040C
03/11/2021	Water	M1647-15	pH	Cool 4 deg C	PSEG03	H51	41145	03/11/2021	9040C
03/11/2021	Water	M1647-16	pH	Cool 4 deg C	PSEG03	H51	40989	03/11/2021	9040C
03/11/2021	Water	M1648-01	pH	Cool 4 deg C	PSEG03	H42	41727	03/11/2021	9040C
03/11/2021	Water	M1648-02	pH	Cool 4 deg C	PSEG03	H42	41728	03/11/2021	9040C
03/11/2021	Water	M1648-03	pH	Cool 4 deg C	PSEG03	H42	30802	03/11/2021	9040C
03/11/2021	Water	M1648-04	pH	Cool 4 deg C	PSEG03	H42	30803	03/11/2021	9040C
03/11/2021	Water	M1648-05	pH	Cool 4 deg C	PSEG03	H42	30804	03/11/2021	9040C
03/11/2021	Water	M1648-06	pH	Cool 4 deg C	PSEG03	H42	30805	03/11/2021	9040C

Date/Time 03/11/2021 15:40
 Received by: JP/604 (Jen)
 Relinquished by: JT (sm)

Date/Time 03/11/2021 17:25
 Received by: JT (sm)
 Relinquished by: JP/604 (Jen)