



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113486

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	03/12/2021	11:25
2	M1647-15		1	10	100.00	>212.0	>212.0	03/12/2021	11:50
3	M1647-15DUP		1	10	100.00	>212.0	>212.0	03/12/2021	12:15
4	M1647-16		1	11	100.00	>212.0	>212.0	03/12/2021	12:40
5	M1647-18		1	12	100.00	>212.0	>212.0	03/12/2021	13:05
6	M1647-19		1	11	100.00	>212.0	>212.0	03/12/2021	13:35
7	M1647-20		1	11	100.00	>212.0	>212.0	03/12/2021	13:58
8	M1647-21		1	10	100.00	>212.0	>212.0	03/12/2021	14:20
9	M1647-22		1	11	100.00	>212.0	>212.0	03/12/2021	14:45
10	M1648-07		1	10	100.00	>212.0	>212.0	03/12/2021	15:10
11	M1648-09		1	11	100.00	>212.0	>212.0	03/12/2021	15:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp31221

WorkList ID : 147098

Department : Wet-Chemistry

Date : 03-12-2021 09:13:09

LB113486

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/18/2021	Water	M1647-15	Flash Point	Cool 4 deg C	PSEG03	H51	41145	03/11/2021	1010A
03/18/2021	Water	M1647-16	Flash Point	Cool 4 deg C	PSEG03	H51	40989	03/11/2021	1010A
03/18/2021	Water	M1647-18	Flash Point	Cool 4 deg C	PSEG03	H51	COMP1-41730:31	03/11/2021	1010A
03/18/2021	Water	M1647-19	Flash Point	Cool 4 deg C	PSEG03	H51	COMP2-41732:34	03/11/2021	1010A
03/18/2021	Water	M1647-20	Flash Point	Cool 4 deg C	PSEG03	H51	COMP4-41735-30780	03/11/2021	1010A
03/18/2021	Water	M1647-21	Flash Point	Cool 4 deg C	PSEG03	H51	COMP5-30811:15	03/11/2021	1010A
03/18/2021	Water	M1647-22	Flash Point	Cool 4 deg C	PSEG03	H51	COMP3-30788:89	03/11/2021	1010A
03/18/2021	Water	M1648-07	Flash Point	Cool 4 deg C	PSEG03	H42	COMP1-41727:28	03/11/2021	1010A
03/18/2021	Water	M1648-09	Flash Point	Cool 4 deg C	PSEG03	H42	COMP2-30802:05	03/11/2021	1010A

Date/Time 03/12/21 11:50
 Received by: KS (wetchem)
 Relinquished by: JP (wetchem)

Date/Time 03/12/21 15:40
 Received by: JP (wetchem)
 Relinquished by: KS (wetchem)