

LB113491


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4117E



A= 0.0000e+000

B= 2.4178e-004

C= 8.6639e-003

Rho= 0.9998743

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%d
0.0	0.000	0.027	0.027	76	0.000	76					0.01
0.2	0.200	0.218	0.018	867	0.0 %	867					9
2.5	2.500	2.454	-0.046	10114	0.0 %	10114					-2
5.0	5.000	4.930	-0.070	20356	0.0 %	20356					-1
7.5	7.500	7.606	0.106	31423	0.0 %	31423					1
10.0	10.000	9.964	-0.036	41178	0.0 %	41178					0

LB113491 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Mar 2021 14:01:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	76	-	0.00007470A	PPB		12 Mar 2021 14:02:39	S	Std
0.2 - 1	SD-2	867	-	0.20007470A	PPB		12 Mar 2021 14:04:55	S	Std
2.5 - 1	SD-5	10114	-	2.50007470A	PPB		12 Mar 2021 14:07:12	S	Std
5.0 - 1	SS	20356	-	5.00007470A	PPB		12 Mar 2021 14:09:29	S	Std
7.5 - 1	SD-5	31423	-	7.50007470A	PPB		12 Mar 2021 14:11:45	S	Std
10.0 - 1	SO	41178	-	10.00007470A	PPB		12 Mar 2021 14:16:37	S	Std
ICV36 - 1	ICV36	16356	3.9631	-7470A	PPB		12 Mar 2021 14:19:13	U	SMPL
ICB36 - 1	ICB36	-30	0.0014	-7470A	PPB		12 Mar 2021 14:21:28	U	SMPL
CCV16 - 1	CCV16	20165	4.8841	-7470A	PPB		12 Mar 2021 14:23:43	U	SMPL
CCB16 - 1	CCB16	-31	0.0012	-7470A	PPB		12 Mar 2021 14:25:58	U	SMPL
CRA - 1	CRA	858	0.2161	-7470A	PPB		12 Mar 2021 14:28:13	U	SMPL
HighStd - 1	HighStd	41349	10.0058	-7470A	PPB		12 Mar 2021 14:30:27	U	SMPL
ChkStd - 1	ChkStd	29629	7.1722	-7470A	PPB		12 Mar 2021 14:32:42	U	SMPL
PB135138BL - 1	PBW	-50	-0.0034	-7470A	PPB		12 Mar 2021 14:34:57	U	SMPL
PB135138BS - 1	LCSW	15861	3.8435	-7470A	PPB		12 Mar 2021 14:37:12	U	SMPL
M1639-02 - 1	STOCKPILE	-9	0.0065	-7470A	PPB		12 Mar 2021 14:39:29	U	SMPL
M1639-02DUP - 1	STOCKPILEDUP	27	0.0152	-7470A	PPB		12 Mar 2021 14:41:45	U	SMPL
M1639-02MS - 1	STOCKPILEMS	16254	3.9385	-7470A	PPB		12 Mar 2021 14:44:02	U	SMPL
M1639-02MSD - 1	STOCKPILEMSD	16521	4.0030	-7470A	PPB		12 Mar 2021 14:46:18	U	SMPL
M1646-01 - 1	RO-310	1	0.0089	-7470A	PPB		12 Mar 2021 14:48:34	U	SMPL
CCV17 - 1	CCV17	20701	5.0137	-7470A	PPB		12 Mar 2021 14:50:49	U	SMPL
CCB17 - 1	CCB17	-20	0.0038	-7470A	PPB		12 Mar 2021 14:53:04	U	SMPL
PB135139BL - 1	PBW	55	0.0220	-7470A	PPB		12 Mar 2021 14:55:19	U	SMPL
PB135139BS - 1	LCSW	15258	3.6977	-7470A	PPB		12 Mar 2021 14:57:34	U	SMPL
M1595-14 - 1	FB02-20210305	4	0.0096	-7470A	PPB		12 Mar 2021 14:59:48	U	SMPL
M1648-07 - 1	COMP1-41727:28	1242	0.3089	-7470A	PPB		12 Mar 2021 15:02:04	U	SMPL
M1648-09 - 1	COMP2-30802:05	156	0.0464	-7470A	PPB		12 Mar 2021 15:04:19	U	SMPL
M1648-09DUP - 1	COMP2-30802:05DUP	113	0.0360	-7470A	PPB		12 Mar 2021 15:06:35	U	SMPL
M1648-09MS - 1	COMP2-30802:05MS	16639	4.0316	-7470A	PPB		12 Mar 2021 15:08:52	U	SMPL
M1648-09MSD - 1	COMP2-30802:05MSD	15224	3.6895	-7470A	PPB		12 Mar 2021 15:11:08	U	SMPL
PB135138TB - 1	PB135124TB	-136	-0.0242	-7470A	PPB		12 Mar 2021 15:13:25	U	SMPL
CCV18 - 1	CCV18	19945	4.8309	-7470A	PPB		12 Mar 2021 15:15:41	U	SMPL
CCB18 - 1	CCB18	-60	-0.0058	-7470A	PPB		12 Mar 2021 15:17:56	U	SMPL
M1639-02LX5 - 1		2	0.0091	-7470A	PPB		12 Mar 2021 15:20:12	U	SMPL
M1639-02A - 1		16397	3.9731	-7470A	PPB		12 Mar 2021 15:22:27	U	SMPL
M1648-09LX5 - 1		-5	0.0075	-7470A	PPB		12 Mar 2021 15:24:42	U	SMPL
M1648-09A - 1		16194	3.9240	-7470A	PPB		12 Mar 2021 15:26:58	U	SMPL
CCV19 - 1	CCV19	20046	4.8553	-7470A	PPB		12 Mar 2021 15:31:28	U	SMPL
CCB19 - 1	CCB19	-29	0.0017	-7470A	PPB		12 Mar 2021 15:33:45	U	SMPL