



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113494

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	03/12/2021	13:45
2	CAL2	1	Water	NA	NA	21.2	7.00	03/12/2021	13:46
3	CAL3	1	Water	NA	NA	21.2	10.03	03/12/2021	13:47
4	ICV	1	Water	NA	NA	21.2	6.99	03/12/2021	13:48
5	CCV1	1	Water	NA	NA	21.1	2.01	03/12/2021	13:50
6	M1677-01	1	Water	NA	NA	20.2	12.60	03/12/2021	14:00
7	M1677-01DUP	1	Water	NA	NA	20.3	12.61	03/12/2021	14:01
8	CCV2	1	Water	NA	NA	21.2	12.02	03/12/2021	14:02

WORKLIST(Hardcopy Internal Chain)

JB113494

WorkList Name : PH M1677 WorkList ID : 147115 Department : Wet-Chemistry Date : 03-12-2021 14:19:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2021	Water	M1677-01'	pH	Cool 4 deg C	TULL02	J42	D2470-D2800(WATER)	03/12/2021	9040C

Date/Time 03/12/21 13:40
 Received by: JP (wet Chem)
 Relinquished by: KS (wet Chem)

Date/Time 03/12/21 15:00
 Received by: KS (wet Chem)
 Relinquished by: JP (wet Chem) Chem