

7471B


LB113498

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56216



A= 0.0000e+000

B= 1.7837e-004

C= -9.1869e-003

Rho= 0.9999561

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.010	0.010	108	0.000	108					
0.2	0.200	0.189	-0.011	1111	0.0 %	1111					-6
2.5	2.500	2.483	-0.017	13974	0.0 %	13974					-1
5.0	5.000	5.056	0.056	28398	0.0 %	28398					-1
7.5	7.500	7.443	-0.057	41782	0.0 %	41782					-1
10.0	10.000	10.018	0.018	56216	0.0 %	56216					0

LB113498 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Mar 2021 15:40:31

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	108	-	0.0000	7471B	PPB	12 Mar 2021 15:41:35	S	Std
0.2 - 1	50.2	1111	-	0.2000	7471B	PPB	12 Mar 2021 15:43:52	S	Std
2.5 - 1	52.5	13974	-	2.5000	7471B	PPB	12 Mar 2021 15:46:10	S	Std
5.0 - 1	55	28398	-	5.0000	7471B	PPB	12 Mar 2021 15:48:26	S	Std
7.5 - 1	57.5	41782	-	7.5000	7471B	PPB	12 Mar 2021 15:50:43	S	Std
10.0 - 1	510	56216	-	10.0000	7471B	PPB	12 Mar 2021 15:52:58	S	Std
ICV37 - 1	ICV37	22343	3.9761	-	7471B	PPB	12 Mar 2021 15:55:37	U	SMPL
ICB37 - 1	ICB37	-22	-0.0131	-	7471B	PPB	12 Mar 2021 15:57:52	U	SMPL
CCV20 - 1	CCV20	28584	5.0893	-	7471B	PPB	12 Mar 2021 16:00:07	U	SMPL
CCB20 - 1	CCB20	-31	-0.0147	-	7471B	PPB	12 Mar 2021 16:02:22	U	SMPL
CRA - 1	CRA	1102	0.1874	-	7471B	PPB	12 Mar 2021 16:04:37	U	SMPL
HighStd - 1	HighStd	56491	10.0671	-	7471B	PPB	12 Mar 2021 16:06:51	U	SMPL
ChkStd - 1	ChkStd	39626	7.0589	-	7471B	PPB	12 Mar 2021 16:09:07	U	SMPL
PB135137BL - 1	PBS	-61	-0.0201	-	7471B	PPB	12 Mar 2021 16:11:24	U	SMPL
PB135137BS - 1	LCSS	22096	3.9321	-	7471B	PPB	12 Mar 2021 16:13:40	U	SMPL
M1625-01 - 1	SP-1-1	358	0.0547	-	7471B	PPB	12 Mar 2021 16:15:56	U	SMPL
M1625-01DUP - 1	SP-1-1DUP	304	0.0450	-	7471B	PPB	12 Mar 2021 16:18:13	U	SMPL
M1625-01MS - 1	SP-1-1MS	21957	3.9073	-	7471B	PPB	12 Mar 2021 16:20:29	U	SMPL
M1625-01MSD - 1	SP-1-1MSD	22040	3.9221	-	7471B	PPB	12 Mar 2021 16:22:46	U	SMPL
M1625-03 - 1	SP-1-2	373	0.0573	-	7471B	PPB	12 Mar 2021 16:25:02	U	SMPL
CCV21 - 1	CCV21	28593	5.0909	-	7471B	PPB	12 Mar 2021 16:27:17	U	SMPL
CCB21 - 1	CCB21	-35	-0.0154	-	7471B	PPB	12 Mar 2021 16:29:33	U	SMPL
M1639-01 - 1	STOCKPILE	23750	4.2271	-	7471B	PPB	12 Mar 2021 16:31:48	U	SMPL
M1640-01 - 1	TP-20-A	10249	1.8189	-	7471B	PPB	12 Mar 2021 16:34:03	U	SMPL
M1649-02 - 1	SP-1-1	537	0.0866	-	7471B	PPB	12 Mar 2021 16:36:18	U	SMPL
M1649-04 - 1	SP-1-2	240	0.0336	-	7471B	PPB	12 Mar 2021 16:38:33	U	SMPL
M1649-06 - 1	SP-1-3	282	0.0411	-	7471B	PPB	12 Mar 2021 16:40:49	U	SMPL
M1650-01 - 1	EO-01-031121	2451	0.4280	-	7471B	PPB	12 Mar 2021 16:43:06	U	SMPL
M1625-01LX5 - 1		123	0.0128	-	7471B	PPB	12 Mar 2021 16:45:22	U	SMPL
M1625-01A - 1		23935	4.2601	-	7471B	PPB	12 Mar 2021 16:47:39	U	SMPL
CCV22 - 1	CCV22	28671	5.1048	-	7471B	PPB	12 Mar 2021 16:49:56	U	SMPL
CCB22 - 1	CCB22	-33	-0.0151	-	7471B	PPB	12 Mar 2021 16:52:11	U	SMPL