



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Alexander

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB113502

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	03/12/2021	16:09
2	CAL2	1	Water	NA	NA	21.1	7.01	03/12/2021	16:11
3	CAL3	1	Water	NA	NA	21.2	10.03	03/12/2021	16:13
4	ICV	1	Water	NA	NA	21.3	7.01	03/12/2021	16:15
5	CCV1	1	Water	NA	NA	21.3	2.02	03/12/2021	16:17
6	M1669-01	1	Solid	20.08	20	21.9	7.46	03/12/2021	16:20
7	M1678-01	1	Solid	20.07	20	21.0	12.45	03/12/2021	16:23
8	M1678-01DUP	1	Solid	20.02	20	20.9	12.42	03/12/2021	16:26
9	M1678-03	1	Solid	20.05	20	20.7	12.44	03/12/2021	16:30
10	CCV2	1	Water	NA	NA	21.3	12.02	03/12/2021	16:32

WORKLIST(Hardcopy Internal Chain)

VB 113 502

WorkList Name : PH M1669 S WorkList ID : 147118 Department : Wet-Chemistry Date : 03-12-2021 15:03:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2021	Solid	M1669-01	pH	Cool 4 deg C	PSEG03	H52	RBR-607	03/12/2021	9045D
03/12/2021	Solid	M1678-01	pH	Cool 4 deg C	TULL02	J41	D2470-D2800(SOIL)-A	03/12/2021	9045D
03/12/2021	Solid	M1678-03	pH	Cool 4 deg C	TULL02	J41	D2470-D2800(SOIL)-B	03/12/2021	9045D

Date/Time 03/12/21 16:05
 Received by: JD (not Open)
 Relinquished by: KS (Welch)

Date/Time 03/12/21 17:05
 Received by: KS (Welch)
 Relinquished by: JD (not Open)