



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB113512

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP89962
Turbidity Calibration std, 0NTU	WP89961
Turbidity Calibration std, 20NTU	WP89957
Turbidity Calibration std, 40NTU	WP89956
Turbidity Calibration std, 1NTU	WP89960
Turbidity Calibration std, 5NTU	WP89959
Turbidity LOQ std, 1.0NTU	WP89963
Turbidity Calibration - CCV std, 10 NTU	WP89958
10 NTU Standard 500 ml	W2752

Intercept: 0.2316

Slope: 0.9864

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.277	0.05		03/15/2021	09:35
2	CAL2	1	1	1.170	0.95	4.9	03/15/2021	09:37
3	CAL3	5	1	5.110	4.95	1.1	03/15/2021	09:39
4	CAL4	10	1	9.800	9.70	3	03/15/2021	09:42
5	CAL5	20	1	20.500	20.55	-2.7	03/15/2021	09:44
6	CAL6	40	1	39.500	39.81	0.5	03/15/2021	09:47



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB113512

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.300	10.207	03/15/2021	09:49
2	ICB		1	0.294	0.063	03/15/2021	09:52
3	CCV1	10	1	9.990	9.893	03/15/2021	09:54
4	CCB1		1	0.287	0.056	03/15/2021	09:57
5	LB113512BL		1	0.301	0.070	03/15/2021	09:59
6	M1458-10		1	0.805	0.581	03/15/2021	10:02
7	M1458-11		1	1.150	0.931	03/15/2021	10:04
8	CCV2	10	1	10.100	10.004	03/15/2021	10:06
9	CCB2		1	0.290	0.059	03/15/2021	10:09