

LB113520


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39208



A= 0.0000e+000

B= 2.5265e-004

C= -1.4360e-002

Rho= 0.9997620

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.005	0.005	78	0.000	78					
0.05	0.050	0.045	-0.005	235	0.0 %	235	0				-10
0.2	0.200	0.184	-0.016	785	0.0 %	785					-8
2.5	2.500	2.408	-0.092	9588	0.0 %	9588					-4
5.0	5.000	5.139	0.139	20399	0.0 %	20399					3
7.5	7.500	7.577	0.077	30047	0.0 %	30047					1
10.0	10.000	9.891	-0.109	39208	0.0 %	39208					-1

LB113520 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Mar 2021 15:14:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	78	-	0.0000	245.1	PPB	15 Mar 2021 15:16:17	S	Std
0.05 - 1	50.05	235	-	0.0500	245.1	PPB	15 Mar 2021 15:18:33	S	Std
0.2 - 1	50.2	785	-	0.2000	245.1	PPB	15 Mar 2021 15:20:50	S	Std
2.5 - 1	52.5	9588	-	2.5000	245.1	PPB	15 Mar 2021 15:23:07	S	Std
5.0 - 1	55	20399	-	5.0000	245.1	PPB	15 Mar 2021 15:25:23	S	Std
7.5 - 1	57.5	30047	-	7.5000	245.1	PPB	15 Mar 2021 15:27:39	S	Std
10.0 - 1	510	39208	-	10.0000	245.1	PPB	15 Mar 2021 15:32:32	S	Std
ICV39 - 1	ICV39	15898	4.0022	-245.1		PPB	15 Mar 2021 15:37:59	U	SMPL
ICB39 - 1	ICB39	-26	-0.0209	-245.1		PPB	15 Mar 2021 15:40:13	U	SMPL
CCV26 - 1	CCV26	20433	5.1480	-245.1		PPB	15 Mar 2021 15:42:28	U	SMPL
CCB26 - 1	CCB26	-34	-0.0229	-245.1		PPB	15 Mar 2021 15:44:43	U	SMPL
CRA - 1	CRA	225	0.0425	-245.1		PPB	15 Mar 2021 15:52:17	U	SMPL
HighStd - 1	HighStd	39682	10.0111	-245.1		PPB	15 Mar 2021 15:54:31	U	SMPL
ChkStd - 1	ChkStd	28125	7.0913	-245.1		PPB	15 Mar 2021 15:56:46	U	SMPL
PB135167BL - 1	PBW	-50	-0.0270	-245.1		PPB	15 Mar 2021 15:59:02	U	SMPL
PB135167BS - 1	LCSW	16180	4.0735	-245.1		PPB	15 Mar 2021 16:01:19	U	SMPL
M1643-01 - 1	32220008-3	-24	-0.0204	-245.1		PPB	15 Mar 2021 16:03:35	U	SMPL
M1651-01 - 1	15220012-1	32	-0.0063	-245.1		PPB	15 Mar 2021 16:05:51	U	SMPL
M1677-01 - 1	D2470-D2800(WATER)	135	0.0197	-245.1		PPB	15 Mar 2021 16:08:07	U	SMPL
M1677-02 - 1	D2470-D2800(WATER)	102	0.0114	-245.1		PPB	15 Mar 2021 16:10:24	U	SMPL
CCV27 - 1	CCV27	20094	5.0623	-245.1		PPB	15 Mar 2021 16:12:39	U	SMPL
CCB27 - 1	CCB27	-17	-0.0187	-245.1		PPB	15 Mar 2021 16:14:54	U	SMPL
M1677-03 - 1	D2470-D2800(WATER)	196	0.0352	-245.1		PPB	15 Mar 2021 16:17:09	U	SMPL
M1677-03DUP - 1	D2470-D2800(WATER)DUP	211	0.0389	-245.1		PPB	15 Mar 2021 16:19:25	U	SMPL
M1677-03MS - 1	D2470-D2800(WATER)MS	15005	3.7766	-245.1		PPB	15 Mar 2021 16:21:40	U	SMPL
M1677-03MSD - 1	D2470-D2800(WATER)MSD	15106	3.8021	-245.1		PPB	15 Mar 2021 16:23:55	U	SMPL
M1677-03LX5 - 1		0	-0.0144	-245.1		PPB	15 Mar 2021 16:26:11	U	SMPL
M1677-03A - 1		16260	4.0937	-245.1		PPB	15 Mar 2021 16:28:26	U	SMPL
CCV28 - 1	CCV28	20367	5.1313	-245.1		PPB	15 Mar 2021 16:30:42	U	SMPL
CCB28 - 1	CCB28	-27	-0.0212	-245.1		PPB	15 Mar 2021 16:32:58	U	SMPL